



Republic of the Philippines
Department of Health
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATION

TWO (2) BRAND NEW ROTARY MICROTOMES
(Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a)(ii) and/or **GCC** Clause 2.1(a)(ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

TWO (2) BRAND NEW ROTARY MICROTOMES

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non- completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>TWO (2) BRAND NEW ROTARY MICROTOMES</u>	PROCUREMENT OF TWO (2) BRAND NEW ROTARY MICROTOMES Fully motorized and automated microtome With manual selection applications Allows multi-sectioning modes including single stroke, continuous stroke, step and rocking modes Manual and motorized sectioning specimen retraction system Coarse feed speed and direction buttons Specimen head movement speed with 2 or more speed settings Adjustable sectioning speed even when sectioning is in progress With knife holder for low profile disposable blades with lateral displacement feature With smooth running handwheel With precision micrometer feed system User friendly integrated or separate control panel With LED display indicators for the following: Specimen retraction Emergency stop Specimen head and hand wheel locks		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) BRAND NEW ROTARY MICROTOMES

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non- completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>TWO (2) BRAND NEW ROTARY MICROTOMES</u>	<p>Emergency stop</p> <p>Specimen head and handwheel locks</p> <p>Trimming and sectioning thickness</p> <p>Retraction status</p> <p>Dual independent handwheel locking system</p> <p>Optical and auditory signals for front and rear travel limits and horizontal feed</p> <p>With microtome base plate ruler</p> <p>With large volume section waste tray</p> <p>With step function in sectioning mode</p> <p>With quick clamping system for specimen holders</p> <p>Sectioning thickness settings of both trimming and sectioning can be save independently</p> <p>With one (1) set Waterbath and Hot Plate</p> <p>With one (1) unit 1500W Auto Voltage Regulator (AVR)</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) BRAND NEW ROTARY MICROTOMES

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non- completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>TWO (2) BRAND NEW ROTARY MICROTOMES</u>	Other terms & conditions of acceptability: 1. The machine must be provided with Uninterrupted Power Supply compatible with the machine with power surge protection. 2. All machines must be manufactured by a known and reputable company with Certificate of Good Manufacturing Practice (GMP), TUV or ISO or its equivalent for equipment only. 3. Supplier must provide company response within 24- 48 hours should the existing equipment breaks down. 4. Supplier must provide preventive maintenance/ calibration of machine as warranted without cost to the hospital.		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) BRAND NEW ROTARY MICROTOMES

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non- completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>TWO (2) BRAND NEW ROTARY MICROTOMES</u>	<p>5. The Supplier must provide the following:</p> <ul style="list-style-type: none">a. Certificate of Calibration with computed values and validity period.b. Three (3) copies of the Operating and Service Manuals in English. <p>6. Supplier must either be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines.</p> <p>7. Supplier must guarantee the following:</p> <ul style="list-style-type: none">a. To have a technical or maintenance personnel based in Davao City.b. To provide 24 hours/day and 7 days/week service support for emergency calls.c. To provide a substitute service engineer and/or specialist in cases where the assigned technical person is out of town.		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) BRAND NEW ROTARY MICROTOMES

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non- completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>TWO (2) BRAND NEW ROTARY MICROTOMES</u>	<p>8. The supplier must bind itself to service the equipment within the warranty period of two (2) years and shall:</p> <ul style="list-style-type: none">a. Conduct monthly preventive maintenance) at its own expense including labor and cost of spare parts.b. Conduct Calibration every six (6) months within the warranty period. <p>9. Supplier's technical personnel shall train the designated operator(s) of the machine and biomedical technician onsite, until they can handle the equipment confidently with no additional cost to the Procuring Entity.</p> <p>10. In case of machine downtime, the supplier shall:</p> <ul style="list-style-type: none">a. Do remedial action (repair and or replacement of spare parts) within 2 days.b. On the 3rd day, if the machine is not operational, a penalty shall be imposed equivalent to the average daily income of the hospital derived from the affected machine.		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) BRAND NEW ROTARY MICROTOMES

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non- completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>TWO (2) BRAND NEW ROTARY MICROTOMES</u>	<p>11. Supplier must upgrade the provided equipment when warranted.</p> <p>12. The end-user and BIOMEDICAL TECHNICIAN must have an actual inspection and or witness an actual installation and commissioning of the machine being offered.</p> <p>13. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in this technical specification and other reasons as stated in the GCC.</p>		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

Name and Signature of Authorized Representative