



Republic of the Philippines  
DEPARTMENT OF HEALTH  
Center for Health Development for Davao Region  
Southern Philippines Medical Center  
J. P. Laurel Avenue, Davao City  
Tel. Nos. 227-27-31; Fax No. 221-7029



## BID BULLETIN NO. 1

(August 2, 2019)

### One (1) Unit Tandem Mass Spectrometry with Ultraperformance Liquid Chromatography (UPLC) Sample Organizer and MSMS Workstation for Expanded Newborn Screening (MS/MS Machine)

Bid No: 2019-07-0386 (ABC NO. 2019-018)

This BID BULLETIN No. 1 is issued to modify or revise some of the *Eligibility and Technical Requirements and Technical Specifications* of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

#### 1. Eligibility and Technical Requirements


- No. 18 requirement has been revised to read *"Notarized Certificate that the bidder shall provide **at least TWO (2) YEARS WARRANTY** on the equipment, parts and service being offered"*.

#### 2. Technical Specifications:

- In page 9, item 5.d, change the *three (3) days* to *five (5) days* for the remedial action and replace the *4th day* to *6th day* for the back-up machine. The requirement now reads *"In case of machine downtime, the supplier is given five (5) days for remedial action (repair and or replacement of spare parts). On the 6th day, if the unit is not operational, a back-up machine shall be provided within the warranty period"*.

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.

  
ATTY. DANILO A. CULLO

Chairman, Bids and Awards Committee