



Republic of the Philippines
DEPARTMENT OF HEALTH
Center for Health Development for Davao Region
Southern Philippines Medical Center
J. P. Laurel Avenue, Davao City
Tel. Nos. 227-27-31; Fax No. 221-7029



BID BULLETIN NO. 1

(March 12, 2019)

**One (1) Unit Brand New Pulsed Dye Laser,
Lease to own for One (1) Year
Bid No: 2019-02-090 (ABC NO. 2019-001)**


This BID BULLETIN No. 1 is issued to modify or revise some items in the *Technical Specifications* of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

1. Technical Specifications

- a. Under the Other Terms and Conditions item 5.d, the remedial period of three (3) days will be changed to five (5) days and the requirements becomes "*In case of machine downtime, the supplier is given **five (5) days** for remedial action (repair and or replacement of spare parts). On the **6th day**, if the unit is not operational, a back-up machine shall be provided (within the warranty period)*".
- b. Under the Other Terms and Conditions item 7, replace the *Automatic Voltage Regulator* to **UPS (at least 150% up KVA)** and the requirement becomes "*The winning bidder shall provide an UPS (at least 150% up KVA) suitable for the equipment*".

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.


LANI P. PALER, MPA, MBA-HA, FPCHA
Chairperson, Bids and Awards Committee