



Republic of the Philippines
DEPARTMENT OF HEALTH
Center for Health Development Davao Region
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



BID BULLETIN NO. 1

(December 14, 2016)

One (1) Brand New Real Time Polymerase Chain Reaction Detection System
Bid No: 2016-11-0702 (ABC NO. 2016-046)

This BID BULLETIN No. 1 is issued to modify or revise some Eligibility and Technical Requirements indicated in the *Checklist of Requirements, Bid Data Sheet (BDS)* and *some items in the Technical Specifications* of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

a. Checklist of Requirements and Bid Data Sheet:

- The required *"Certification that the equipment being offered must be of the latest version or model and has no record of violations such as electrical or mechanical malfunction since the time of its production"* must be issued by the principal.
- The Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, *whether similar or not similar in nature and complexity to the contract to be bid* shall be supported by the **Purchase Order or Contract Agreement or NOA** stating the amount of the project;

b. Technical Specifications:

- In page 2, there was a revision in the specification for CALIBRATED DYES and it must include FAM, SYBR®Green, HEX,VIC, **CY3 or its equivalent**, ROX and other commonly used dyes.
- The required ten (10) years availability of spare parts will be lowered to five years from the warranty period and the requirement becomes ***"Notarized certificate of the availability of spare parts for the next five (5) years from the warranty period"***.

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.


RICARDO S.D. JUSTOL, MARE, MPA
Chairperson, Bids and Awards Committee