



Republic of the Philippines  
DEPARTMENT OF HEALTH  
Center for Health Development Davao Region  
**SOUTHERN PHILIPPINES MEDICAL CENTER**  
J.P. Laurel Avenue, Davao City



## **BID BULLETIN NO. 1**

(December 11, 2015)

### **PROSTATE TRIPLANE TRANSDUCER COMPATIBLE WITH EXISTING ULTRASOUND SCANNER**

Bid No: 2015-11-0490(ABC NO. 2015-080)

This BID BULLETIN No. 1 is issued to modify or revise some items of the **Technical Specifications** of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

1. In page 3, delete the following items:  
*With one (1) leakage testing kit*  
*With one (1) magnetic wheel mover for 3D imaging*  
*With one (1) mover adaptor ring*
2. In page 4:
  - a. Delete the *one (1) reusable* and the item becomes “*With one (1) reusable endfire biopsy guides, bore diameter of 1.6mm, angle of insertion of 19 degrees to the long axis of the transducer with dual biopsy guides, bore diameter of 1.6mm, angle of insertion of 0 degrees to the long axis of the transducer*”.
  - b. Branded computer set accessories includes computer table, chair, laserjet colored printer, speaker and web camera.
3. Under the Other Terms and Conditions of Acceptability, the required *Service Engineer or Technical personnel based in Davao City* will no longer be required as long that the company’s *Service Engineer or Technical personnel will response within 24-48 hours to answer client calls in case of technical problems or concerns and to do preventive maintenance based on the required schedule.*

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.

  
**RICARDO S.D. JUSTOL, MARE, MPA**  
Chairperson, Bids and Awards Committee