



Republic of the Philippines
DEPARTMENT OF HEALTH
Center for Health Development Davao Region
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



BID BULLETIN NO. 1

(December 11, 2015)

SLIT LAMP PACKAGE

Bid No: 2015-11-0480 (ABC NO. 2015-080)

This BID BULLETIN No. 1 is issued to modify or revise some items of the **Technical Specifications** of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

1. In page 2
 - a. Item 3, delete adjustment range -5 to +5D and replace -8 to +8D. The item will become *“Eyepiece diopter adjustment range from -8 to +8D”*.
 - b. Item 13, the item must be *“With Goldmann Applanation tonometer unit”*.
 - c. Add ***“Ophtha Examining Chair*** with the following specifications:
Reclinable
Up-and-down control
Purpose: for chart projector, phoropter and light source
Rotation: 0-90°
 - d. Add *“Upgradable slit lamp with photo video capability”*.

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.


RICARDO S.D. JUSTOL, MARE, MPA
Chairperson, Bids and Awards Committee