



Republic of the Philippines  
DEPARTMENT OF HEALTH  
Center for Health Development for Davao Region  
Southern Philippines Medical Center  
J. P. Laurel Avenue, Davao City  
Tel. Nos. 227-27-31; Fax No. 221-7029



## BID BULLETIN NO. 1

(March 12, 2019)

Supply, Delivery and Installation of Tandem Mass Spectrometry with Ultra Performance Liquid Chromatography (UPLC) Sample Organizer for Expanded Newborn Screening (MS/MS Machine)

Bid No: 2019-02-091 (ABC NO. 2019-003)


This BID BULLETIN No. 1 is issued to modify or revise some items in the *Technical Specifications* of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

### 1. Technical Specifications

- a. Under the Other Terms and Conditions item 7, the remedial period of three (3) days will be changed to five (5) days and the requirements becomes *"In case of machine downtime, the supplier is given **five (5) days** for remedial action (repair and or replacement of spare parts). On the **6<sup>th</sup> day**, if the unit is not operational, a back-up machine shall be provided (within the warranty period)"*.
- b. Under the Other Terms and Conditions item 12, *the winning bidder shall provide an **UPS (at least 150% up KVA)** suitable for the equipment"*.

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.

  
LANI P. PALER, MPA, MBA-HA, FPCHA  
Chairperson, Bids and Awards Committee