



Republic of the Philippines
DEPARTMENT OF HEALTH
Center for Health Development Davao Region
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



BID BULLETIN NO. 1

(September 6, 2016)

One (1) Unit Brand New Premium Ultrasound, One (1) Unit Midrange Ultrasound Machine with 4D Capability & One (1) Unit Midrange Ultrasound without 4D Capability, Lease to Own for Two (2) Years, 1 Lot

Bid No: 2016-08-0340 (ABC NO. 2016-002)

This BID BULLETIN No. 1 is issued to modify or revise some items of the **Technical Specifications** of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

1. The following were some changes in the specifications of One (1) unit Midrange Ultrasound with 4D Capability:

- In page 8
 - Delete *elastography, SonoNT with SonoIT* and the specification becomes "*Capable of anatomical realism using 3D/4D rendering, spatio temporal image correlation, advanced SRI, (tomographic ultrasound imaging, vocal II, inversion*".
 - Delete "*With one (1) Wideband micro convex probe with 4-10 MHz frequency range for ObGyn, Urology*".
- In page 9
 - Delete *convex* in the *One (1) convex 2D probe 2-8 mhz* and the specification becomes "***One (1) linear 2D probe 2-8 mhz***".
 - Delete "*One (1) endovaginal 2D probe*".
 - Delete *USB* in the *With USB, footswitch, digital printer, memory sticks* and the specification becomes "*With footswitch, digital printer, memory sticks*".

2. The following were some changes in the specification of One (1) unit Midrange Ultrasound without 4D Capability:

- In page 11, correction to the word *sweel*, the correct word is *sweep*. The specification becomes *Capable of Doppler audio playback in cine including baseline and **sweep***.
- In page 12, add in the specification "***With one (1) footswitch***"

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.


RICARDO S. JUSTOL, MARE, MPA
Chairperson, Bids and Awards Committee



Republic of the Philippines
DEPARTMENT OF HEALTH
Center for Health Development Davao Region
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



BID BULLETIN NO. 1

(September 6, 2016)

One (1) Unit Brand New Premium Ultrasound, One (1) Unit Midrange Ultrasound Machine with 4D Capability & One (1) Unit Midrange Ultrasound without 4D Capability, Lease to Own for Two (2) Years, 1 Lot

Bid No: 2016-08-0340 (ABC NO. 2016-002)

This BID BULLETIN No. 1 is issued to modify or revise some items of the **Technical Specifications** of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

1. The following were some changes in the specifications of One (1) unit Midrange Ultrasound with 4D Capability:

- In page 8
 - Delete *elastography, SonoNT with SonoIT* and the specification becomes *"Capable of anatomical realism using 3D/4D rendering, spatio temporal image correlation, advanced SRI, (tomographic ultrasound imaging, vocal II, inversion"*.
 - Delete *"With one (1) Wideband micro convex probe with 4-10 MHz frequency range for ObGyn, Urology"*.
- In page 9
 - Delete *convex* in the *One (1) convex 2D probe 2-8 mhz* and the specification becomes **"One (1) linear 2D probe 2-8 mhz"**.
 - Delete *"One (1) endovaginal 2D probe"*.
 - Delete *USB* in the *With USB, footswitch, digital printer, memory sticks* and the specification becomes *"With footswitch, digital printer, memory sticks"*.

2. The following were some changes in the specification of One (1) unit Midrange Ultrasound without 4D Capability:

- In page 11, correction to the word *sweel*, the correct word is *sweep*. The specification becomes *Capable of Doppler audio playback in cine including baseline and **sweep***.
- In page 12, add in the specification **"With one (1) footswitch"**

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.


RICARDO S.D. JUSTOL, MARE, MPA
Chairperson, Bids and Awards Committee



Republic of the Philippines
DEPARTMENT OF HEALTH
Center for Health Development Davao Region
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



BID BULLETIN NO. 1

(September 6, 2016)

One (1) Unit Brand New Premium Ultrasound, One (1) Unit Midrange Ultrasound Machine with 4D Capability & One (1) Unit Midrange Ultrasound without 4D Capability, Lease to Own for Two (2) Years, 1 Lot

Bid No: 2016-08-0340 (ABC NO. 2016-002)

This BID BULLETIN No. 1 is issued to modify or revise some items of the **Technical Specifications** of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

1. The following were some changes in the specifications of One (1) unit Midrange Ultrasound with 4D Capability:

- In page 8
 - Delete *elastography, SonoNT with SonoIT* and the specification becomes *"Capable of anatomical realism using 3D/4D rendering, spatio temporal image correlation, advanced SRI, (tomographic ultrasound imaging, vocal II, inversion"*.
 - Delete *"With one (1) Wideband micro convex probe with 4-10 MHz frequency range for ObGyn, Urology"*.
- In page 9
 - Delete *convex* in the *One (1) convex 2D probe 2-8 mhz* and the specification becomes **"One (1) linear 2D probe 2-8 mhz"**.
 - Delete *"One (1) endovaginal 2D probe"*.
 - Delete *USB* in the *With USB, footswitch, digital printer, memory sticks* and the specification becomes *"With footswitch, digital printer, memory sticks"*.

2. The following were some changes in the specification of One (1) unit Midrange Ultrasound without 4D Capability:

- In page 11, correction to the word *sweel*, the correct word is *sweep*. The specification becomes *Capable of Doppler audio playback in cine including baseline and **sweep***.
- In page 12, add in the specification **"With one (1) footswitch"**

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.


RICARDO S.D. JUSTOL, MARE, MPA
Chairperson, Bids and Awards Committee