



BID BULLETIN NO. 2

(September 2, 2019)


Three (3) Units Brand New Perinatology Digital Color Doppler Ultrasound Bid No: 2019-08-01-026 (ABC NO. 2019-021)

This BID BULLETIN No. 2 is hereby issued to modify or revise the *Technical Specifications* of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

1. The title of the project has been revised to read ***"TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND"***.
2. Clarifications or Revisions to the Technical Specifications:
 - 2.1 In page 2 under the *System Design*
 - Bullet 2, the specification has been revised to read ***"Must have at least 4 active ports and 1 pencil port and at least 5 probe holders"***
 - 2.2 In page 5 under the *Primary Storage*
 - Bullet 1, delete ***"or its equivalent"*** and the specification has been revised to read ***"Linux Operating System, with Hard Drive memory of not less than 1TB"***
 - 2.3 In page 6 under the *Transducers*
 - Bullet 4, delete ***"Wideband Linear Probe (3 to 17MHz). Clinical Applications: Vascular, Small Parts, MSK and Breast"***
 - Add in the specification ***"Endocavity Volumetric Probe (3 to 13MHz). Clinical Applications: Gynecology, Obstetrics, Urology"***
3. Due to many items in the technical specifications that were revised, a revised technical specification shall be issued to prospective bidders who have bought the bidding documents. The revised technical specification must be submitted during the bid opening.

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.


ATTY. DANILO A. CULLO
Chairman, Bids and Awards Committee