



Republic of the Philippines  
DEPARTMENT OF HEALTH  
Center for Health Development for Davao Region  
Southern Philippines Medical Center  
J. P. Laurel Avenue, Davao City  
Tel. Nos. 227-27-31; Fax No. 221-7029



## BID BULLETIN NO. 2

(September 2, 2019)

**Supply, Delivery and Installation of Fifteen (15) Units Brand New High Accuracy Pulse Oximeters**  
Bid No: 2019-08-01-011 (ABC NO. 2019-021)

This BID BULLETIN No. 2 is hereby issued to modify or revise some of the *Technical Specifications* of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

1. Clarifications or Revisions of the Technical Specifications:

1.1 In page 2 under the *Fifteen (15) Units Brand New High Accuracy Pulse Oximeters*

- Bullet 12, the specification for "Pulse rate" has been revised to now read "Pulse rate measurement range of **at least 20 - 250 beats per minute**".

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.



ATTY. DANILO A. CULLO

Chairman, Bids and Awards Committee