



## BID BULLETIN NO. 2

(December 5, 2020)

### 64 Slices Dual Energy CT Scanner System

Bid No: 2020-11-05-090 (ABC NO. 2020-042)

This BID BULLETIN No. 2 is issued to modify or revise some items in the *Instruction to Bidders, Special Conditions of the Contract and Technical Specifications* of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

#### 1. Instruction to Bidders

- Under the Instruction to Bidders Clause 14. 2, change April 13, 2020 to *April 13, 2021* and the statement now read *"The Bid and bid security shall be valid until **April 13, 2021** or one hundred twenty (120) calendar days from the date of the opening of bids"*.

#### 2. Special Conditions of the Contract (under Clause 1 additional technical requirements)

- In page 22 number 11, change the five (5) years warranty to *three (3) years* and the requirement now read *"Notarized Certificate that the bidder shall provide at least **THREE (3) YEARS WARRANTY** on the equipment, parts and service being offered"*.
- In page 23 number 12, change the five (5) years warranty to *three (3) years* and the requirement now read *"Commitment under oath (Notarized) Certification from the manufacturer or principal that any changes of their Philippine exclusive representative will not in any way affect the three (3) years warranty of the equipment. If the certificate is issued from abroad, the signature of its notary public must be authenticated by the Philippine Consulate Office in that country (Apostille). In lieu of the original copy of the document, a photocopy certified as true copy by a practicing lawyer in the Philippines shall be submitted"*.


#### 3. The technical specifications shall be revised to the statement as follows:

- Page 31 under item II. *Gantry*, change not be less than 80 cm to *not be less than 70 cm* and the specification now read *"The Gantry aperture must **not be less than 70 cm**"*.
- Page 32 under item VI. *Patient Couch*, change the 1600 mm (160 cm) to *1580 mm (158 cm)* and the specification now read *"The scan range for axial and helical scanning must be **1580 mm (158 cm)** or higher with table extension"*.

- Page 34 revised the size of the LCD monitors from 20 inches to 19 inches and the specification now read *"The console shall be provided with two (2) colored LCD medical grade monitors with diagonal dimension of more than 19 inches"*.
- Page 39 under the XIX. Accessories, change the at least 24" to at least 23" and the specification now read *"Real time CT Flouroscope Hardware such as footswitch, **at least 23"** in room head ceiling mounted medical grade LCD monitor; exposure control, couch control"*.
- Page 43 paragraph 6 under the XXI. Other Terms and Conditions, change at least five (5) years to at least three (3) years and now read *"The supplier must provide **at least three (3) years** full parts and service warranty of the CT Scanner and its accessories such as multimedia printer, automatic injector, AVR and UPS (the electrical capacity requirement of the UPS is 150% total capacity of the equipment) including its batteries. The warranty shall commence after passing the conformance test conducted by the Center for Device Regulation Radiation Health Research of the Foods and Drugs Administration (FDA-CRRHR)"*.

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.

  
ATTY. DANILO A. CULLO  
Chairman, Bids and Awards Committee