



BID BULLETIN NO. 2

(December 5, 2020)

ONE HUNDRED (100) UNITS BRAND NEW PSYCHIATRY BEDS

Bid No: 2020-11-05-091 (ABC NO. 2019-021)

This BID BULLETIN No. 2 is issued to modify or revise some items in the *Instruction to Bidders, Special Conditions of the Contract and Technical Specifications* of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

1. Instruction to Bidders

- Under the Instruction to Bidders Clause 14. 2, change April 13, 2020 to *April 13, 2021* and the statement now read *"The Bid and bid security shall be valid until **April 13, 2021** or one hundred twenty (120) calendar days from the date of the opening of bids"*.

2. Special Conditions of the Contract

- In page 22 under Clause 1 List of additional technical requirements for the completion of this Contract, revised item no. 3 and item no. 5. The requirements now read:
 - Item No. 3 *"Notarized Certificate that the bidder shall provide at least TWO (2) YEARS WARRANTY on the beds being offered"*.
 - Item No. 5 *"Certification that the bed being offered must be brand new, unused, and was not listed in the product recall"*.


3. Technical Specifications

- In page 29
 - Under the Other Terms and Conditions, revised item no. 1 and now read *"The item must be BRAND-NEW and under warranty for at least 2 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts)"*.
- In page 30 and 31
 - Delete item no. 3 *"The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of SPMC Biomedical Technician, End-user and the Company Engineer"*.
 - Delete letter b to d in item no. 4 and the requirement becomes *"Supplier must be the exclusive or authorized distributor of the principal company of the bed in the Philippines"*.
 - Delete item no. 5, item no. 6 and item no. 7 of the Other Terms and Conditions:

- No. 5 *"The Supplier must submit Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment"*.
- No. 6. *"The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end-users and biomedical technician without additional cost to the procuring entity. Supplier must submit a comprehensive training module as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist"*.
- No. 7 *"Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier"*.

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.


ATTY. DANILO A. CULLO
Chairman, Bids and Awards Committee