



## BID BULLETIN NO. 2

(December 5, 2020)

### ONE (1) UNIT BRAND NEW DIGITAL DENTAL X-RAY

IB NO. 2020-11-05-092 (ABC NO. 2019-021)

This BID BULLETIN No. 2 is issued to modify or revise some items in the *Instruction to Bidders, Special Conditions of the Contract and Technical Specifications* of the abovementioned project. This modification shall prevail and form an integral part of the Bidding Documents.

#### 1. Instruction to Bidders

- Under the Instruction to Bidders Clause 14. 2, change April 13, 2020 to *April 13, 2021* and the statement now read *"The Bid and bid security shall be valid until **April 13, 2021** or one hundred twenty (120) calendar days from the date of the opening of bids"*.

#### 2. Special Conditions of the Contract


- Under Clause 1 List of additional technical requirements for the completion of this Contract, delete No. 6 *"HL7 Conformance Statement for the equipment being offered"*.

#### 3. Technical Specifications

- In page 30
  - Under No. 2 Sensor, the specification has been revised and now read *"Sensor CMOS/CMOS or CMOS/ Flat-Panel Photodiode Array (PANO/CBCT)"*.
  - Under No. 3 Generator, replace the *not less than 50 kv to 99 kv* to **not less than 50 kv to 90 kv** and the specification has been revised and now read *"Tube Voltage: not less than 50 kv to 90 kv"*

All other provision of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.

  
ATTY. DANILO A. CULLO  
Chairman, Bids and Awards Committee