



Republic of the Philippines
Department of Health
Regional Office No. XI
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

(Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a) (ii) and/or **GCC** Clause 2.1(a) (ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND</u></p>	<p>PROCUREMENT OF TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND</p> <p>SYSTEM DESIGN:</p> <ul style="list-style-type: none"> ○ The Ultrasound system must have B, Simultaneous Dual, M, Color M, Anatomical M-mode, Doppler (Color, Spectral and Power) and triplex modes. ○ Must have at least 4 active ports and 1 pencil port and at least 5 probe holders ○ Must have built-in/integrated CD-R/DVD-R and at least 6 USB ports 2 of which should be accessibly located in front of the unit. ○ Must have full DICOM capability with LAN connectivity and WLAN capability. ○ Must have at least 21.5" LED articulating Monitor at minimum resolution of 1920x1080 pixels with Anti-Flickering technology and manual adjustment for Contrast and Brightness. ○ Must have at least 13.3" User Configurable Touch Screen Control Panel and Physical QWERTY Keyboard. ○ Must have built-in Gel Warmer. ○ Must have Rear and Front Handles for easy transport. ○ Must have motorized control panel height adjustment with built-in Physical QWERTY keyboard 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>TWO (2) UNITS</u> <u>BRAND NEW</u> <u>PERINATOLOGY</u> <u>DIGITAL COLOR</u> <u>DOPPLER</u> <u>ULTRASOUND</u></p> <p><i>(continuation)</i></p>	<p>IMAGING TECHNOLOGY:</p> <p>The Ultrasound Machine should have these following Imaging technologies:</p> <ul style="list-style-type: none"> ○ Tissue Harmonic Technologies and Phase Inversion Harmonic Imaging ○ Spatial Compounding Technology ○ Frequency Compounding Technology ○ Speckle Reduction Imaging ○ 3D Speckle Reduction Imaging ○ Automatic 1-click Image optimization for B, Color & Spectral Imaging mode ○ Trapezoidal Imaging ○ 2D and Color Panoramic Imaging with indicator of scanning speed ○ Needle enhancement function for biopsy guide ○ Real-Time 3D and 4D Imaging with Auto Face detection, Auto NT measurement and Multi-Cut which can allow manual adjustment of 4D signals ○ Real-Time 3D and 4D Imaging with at least 8 kinds of 4D rendering mode ○ Includes rendering mode which can allow doctors to get transparent volume images with more comprehensive internal and external view of anatomy structure ○ Includes rendering mode which can allow doctors to identify depth information based on color difference ○ Includes rendering mode which provide tomography slices of 4D images, with slice number up to 29 and slice space with 0.5-10.0mm 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>TWO (2) UNITS</u> <u>BRAND NEW</u> <u>PERINATOLOGY</u> <u>DIGITAL COLOR</u> <u>DOPPLER</u> <u>ULTRASOUND</u></p> <p><u>(continuation)</u></p>	<ul style="list-style-type: none"> ○ Spatio-Temporal Image Correlation (STIC) ○ COLOR 3D imaging for more intuitive and realistic blood flow imaging, indicates clear spatial distribution of vascular networks ○ One-key 2D optimization technology using Artificial Intelligence, which could automatically get the optimistic plane for measurement out of a 2D scanning cine loop, and provide automatic measurement for BPD/HC, AC, FL based on AI deep learning technology. ○ Elastography Imaging for Linear & Convex & TransV probes. ○ Micro Flow Imaging which can provide contrast-enhanced images of small vessels spatial resolution and detail. ○ Micro Flow Imaging which can indicate blood perfusion information with color coded image. <p>IMAGING PARAMETERS</p> <ul style="list-style-type: none"> ○ The Ultrasound Machine must have solid-state designs based on digital broad band technology. ○ Must have a dynamic range of at least 20-280Db, Imaging level of 10-bit and Imaging channel of 128 physical imaging channels. ○ Must have Chroma with not less than 12 types of color ○ Must have up to 11 focus number with zone focusing technology. ○ Must have a minimum of 1500 Frame rates per second (probe dependent). ○ Must allow retrospective manipulation of raw acquisition data. 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>TWO (2) UNITS</u> <u>BRAND NEW</u> <u>PERINATOLOGY</u> <u>DIGITAL COLOR</u> <u>DOPPLER</u> <u>ULTRASOUND</u></p> <p><i>(continuation)</i></p>	<ul style="list-style-type: none"> ○ Must have a maximum depth of 40cm (probe dependent). ○ Manual selection of frequency is required with not less than 5 bands. ○ Must have up to 10x zoom capability. ○ Must have a minimum of 8 TGC adjustment and Lateral Gain adjustments. ○ The system must allow user to insert annotation and anatomical markers. ○ The system must save sonographer specific image presets. <p>MEASUREMENTS AND CALCULATIONS:</p> <ul style="list-style-type: none"> ○ The system must be able to perform 10 caliper measurements on a single image with proper numerical measurement. ○ Must have 3dist Volume measurements, Ellipse, Manual Trace, depth and ratio. ○ Must be able to automatically trace spectral doppler with PI and RI results. ○ Must have complete measurements package for OB/GYN packages for both basic and congenital scans. ○ Automated Volume Count follicle function on 4D endocavity probe. <p>PRIMARY STORAGE</p> <ul style="list-style-type: none"> ○ Linux Operating System with Hard Drive memory of not less than 1TB 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>TWO (2) UNITS</u> <u>BRAND NEW</u> <u>PERINATOLOGY</u> <u>DIGITAL COLOR</u> <u>DOPPLER</u> <u>ULTRASOUND</u></p> <p><i>(continuation)</i></p>	<p>TRANSDUCERS:</p> <ul style="list-style-type: none"> ○ The probes shall include multi frequency imaging capability for optimal resolution and penetration. ○ The Endocavity probe should have Temperature detection technology, with indication of real time temperature of probe head on the monitor. ○ Single crystal Convex Probe (1 to 8Mhz). Clinical Applications: Abdomen, Obstetrics and Gynecology, Urology and Fetal Heart. ○ Endocavity Volumetric Probe (3 to 13MHz). Clinical Applications: Gynecology, Obstetrics, Urology. ○ Endocavity Probe (3 to 15MHz). Clinical Applications: Gynecology, Obstetrics, Urology. With no less than 190 degree FOV and temperature detection technology, can support elastography function for endocavity checking ○ Endocavity Probe (3 to 15 MHz). with compatible Speculum for IVF, operation and other purposes. For continuous monitoring during procedure. Clinical Application Gynecology ○ Wideband Volume Convex Probe (2 to 12MHz). Clinical Applications: Obstetrics, Abdominal and Gynecology ○ Single Crystal Adult Phased Array Probe (1 to 7 MHz). Clinical Applications: Cardiac, Transcranial and Pediatric Abdomen ○ Pediatric Phased Array Probe (2 to 9 MHz). Clinical Applications: Pediatric Echo, Neonatal Echo 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>TWO (2) UNITS</u> <u>BRAND NEW</u> <u>PERINATOLOGY</u> <u>DIGITAL COLOR</u> <u>DOPPLER</u> <u>ULTRASOUND</u></p> <p><i>(continuation)</i></p>	<p>PATIENT DATA MANAGEMENT:</p> <ul style="list-style-type: none"> ○ The system must have built-in patient and report review tool ○ The system must support JPG, BMP and TIF Image format ○ The system must support AVI and WMV Video format ○ The system must support PDF Document format <p>ELECTICAL REQUIREMENTS:</p> <ul style="list-style-type: none"> ○ The system must be at maximum of 120kg without accessories ○ The system must operate at 220-240Volts and Frequency of 50/60Hz. <p>ACCESSORIES:</p> <ul style="list-style-type: none"> ○ Two (2) units Thermal Printer ○ Thee (3) units 2KVA double conversion UPS suitable for the equipment 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>TWO (2) UNITS</u> <u>BRAND NEW</u> <u>PERINATOLOGY</u> <u>DIGITAL COLOR</u> <u>DOPPLER</u> <u>ULTRASOUND</u></p> <p><i>(continuation)</i></p>	<p>Other terms & conditions of acceptability:</p> <ol style="list-style-type: none"> 1. The equipment must be BRAND-NEW and under warranty including spare parts for at least 2 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts). 2. All machines must be manufactured by a known and reputable company compliant with IEC or ISO, or its equivalent for equipment only. 3. The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of a Biomedical Technician, End-user and the Company Engineer. 4. In line with the ISO standard requirement, the engineer must be capable of conducting annual calibration and the issuance of certificate for the said calibration based on the following data: <ol style="list-style-type: none"> a. Reference standards b. National Institute of Standards and Technology traceability c. Reference values d. Validity period e. Uncertainty of measurement 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

Name of Project

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>TWO (2) UNITS</u> <u>BRAND NEW</u> <u>PERINATOLOGY</u> <u>DIGITAL COLOR</u> <u>DOPPLER</u> <u>ULTRASOUND</u></p> <p><i>(continuation)</i></p>	<p>5. Supplier must comply with the following conditions:</p> <ul style="list-style-type: none"> a. Must be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines. b. Submit Certificate of Training of the Company Engineer / Technical Personnel and Product Specialist issued by the Principal or Manufacturer. c. Provide company response within 24 – 48 hours in case of technical problems or equipment breakdown. d. In case of equipment downtime, the supplier is given three (3) days for remedial action (repair and or replacement of spare parts). On the 4th day, if the unit is not operational, a back-up equipment shall be provided (within the warranty period). <p>6. The Supplier must submit the following:</p> <ul style="list-style-type: none"> a. Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment. b. Notarized Certificate of availability of spare parts for the next ten (10) years. 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p>TWO (2) UNITS BRAND NEW PERINATOLOGY DIGITAL COLOR DOPPLER ULTRASOUND</p> <p><i>(continuation)</i></p>	<p>6. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end-users and biomedical technician without additional cost to the procuring entity. Supplier must submit a comprehensive training module as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist.</p> <p>7. The winning bidder must provide and maintain all the necessary computer hardware and the necessary computer accessories, as required.</p> <p>8. Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier.</p> <p>9. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the General Conditions of the Contract (GCC).</p>		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

Name and Signature of Authorized Representative