



Republic of the Philippines
Department of Health
Regional Office No. XI
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATIONS

ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM

(Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a) (ii) and/or **GCC** Clause 2.1(a) (ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM</u>	<u>ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM</u> One (1) Unit Full High Definition Medical Grade Monitor, not less than 27 inches One (1) unit High Definition Video System with Optical Image Enhancement Technology 16:9 and 16:10 output for an HDTV monitor Compatible with analog HD-SDI and DVI output With link connection to peripheral devices for high transmission speed and less cables With Picture-in-picture and index function capability With portable memory for data management Capable of DV output to standard documentation devices Weight of not more than 11 kilograms One (1) unit Xenon Light Source 300 watts Not less than 500 hours lamp life With one link waterproof connector to the light source		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(Continuation)</i> <u>ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM</u>	Scope cable free With link connection to peripheral devices With automatic light adjustment for high quality illumination Weight of not more than 20 kilograms One (1) unit Choledochovideoscope Optical system field of view of 120 degrees forward viewing Depth of field of 3 to 50 millimeters Distal end outer diameter of 4.9 millimeters Bending section angulation range of 160 degrees up and 130 degrees down Working length of 600 millimeters Insertion tube outer diameter of 380 millimeters Total length of 660 millimeters Biopsy forceps minimum visible distance of 3 millimeters from distal end Instrument channel inner diameter of 2.0 millimeters Accessories One (1) piece suction cleaning adapter Ten (10) pieces biopsy valve One (1) piece channel cleaning brush		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(Continuation)</i> <u>ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM</u>	One (1) piece leakage tester One (1) piece biopsy forceps One (1) piece grasping forceps basket type One (1) piece washing pipe Ten (10) pieces suction valve One (1) piece CBD grasping forceps, 330 millimeters, jaw length: 16 millimeters One (1) piece Bipolar Maryland forceps, 330 millimeters, jaw length of 14.7 millimeters One (1) piece Bipolar HF cable One (1) unit videocart One (1) unit AVR One (1) unit medical grade DVD recorder Other terms & conditions of acceptability: 1. The equipment must be a BRAND NEW unit and under warranty including spare parts for at least 2 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts in compliance) with SPMC ISO and CQI Standards.		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM</u>	<p>2. All machines must be manufactured by a known and reputable company with Certificate of Good Manufacturing Practice (GMP), TUV or ISO or its equivalent for equipment only.</p> <p>3. The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of a Biomedical Technician, End-user and the Company Engineer.</p> <p>4. In line with the ISO standard requirement, the engineer must be capable of conducting annual calibration and the issuance of certificate for the said calibration based on the following data:</p> <ul style="list-style-type: none">a. Reference standardsb. National Institute of Standards and Technology traceabilityc. Reference valuesd. Validity periode. Uncertainty of measurement <p>5. Supplier must comply with the following conditions:</p> <ul style="list-style-type: none">a. Must be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines.		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><i>(continuation)</i></p> <p><u>ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM</u></p>	<p>b. Submit Certificate of Training of the Company Engineer / Technical Personnel and Product Specialist issued by the Principal or Manufacturer.</p> <p>c. Provide company response within 24 – 48 hours in case of technical problems or equipment breakdown.</p> <p>d. In case of machine downtime, the supplier is given three (3) days for remedial action (repair and or replacement of spare parts). On the 4th day, if the unit is not operational, a back-up machine shall be provided (within the warranty period).</p> <p>6. The Supplier must submit the following:</p> <p>a. Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment.</p> <p>b. Notarized Certificate of availability of spare parts for the next ten (10) years.</p> <p>7. The winning bidder shall provide an Automatic Voltage Regulator suitable for the equipment.</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>ONE (1) SET BRAND NEW CHOLEDOCHOSCOPY SYSTEM</u>	<p>8. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end-users and biomedical technician without additional cost to the procuring entity. Supplier must submit a comprehensive training module as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist.</p> <p>9. Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier.</p> <p>10. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the General Conditions of the Contract (GCC).</p>		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

Name and Signature of Authorized Representative