



Republic of the Philippines
Department of Health
Regional Office No. XI
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATIONS

FOUR (4) UNITS BRAND NEW ELECTROHYDRAULIC OPERATING TABLES

(Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a) (ii) and/or **GCC** Clause 2.1(a) (ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

FOUR (4) UNITS BRAND NEW ELECTROHYDRAULIC OPERATING TABLES

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>FOUR (4) UNITS BRAND NEW ELECTROHYDRAULIC OPERATING TABLES</u>	<u>FOUR (4) UNITS BRAND NEW ELECTROHYDRAULIC OPERATING TABLES</u> Electromotor adjustment for height, back section, sliding, Trendelenburg and lateral lift Modular type Detachable and light weight table top sections Gas spring assisted divided leg sections with not more than 26-27 inches length Gas spring assisted leg section, 24-25 inches Table top sections with quick release mechanism Table top extension, 10-11 inches Table top length of 82 inches and width of 21 inches Height adjustment range of 23 to 35 inches Lateral tilt capability of -20 to 20 degrees Trendelenburg and reverse Trendelenburg adjustment up to 26 degrees Longitudinal shift adjustment to 11-12 inches Head section adjustment from -45 to 45 degrees Back section adjustment from -40 to 70 degrees Leg section adjustment from -90 to 20 degrees With standard back rest		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

FOUR (4) UNITS BRAND NEW ELECTROHYDRAULIC OPERATING TABLES

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(Continuation)</i> <u>FOUR (4) UNITS</u> <u>BRAND NEW</u> <u>ELECTROHYDRAULIC</u> <u>OPERATING TABLES</u>	With double articulated head rest With 3 inches visco-elastic foam mattress with air filter system Anti static and latex free foam cover With standard seat section with sliding With automated 0-position capability X-ray translucent table top sections With back up control Welded seams Hand control unit with power switch, with ready indicator, height, tilt, Trendelenburg, back section function, sliding function and pedal control With back up control panel in the column battery and charging status indicator Safe working load of not less than 225 kilograms Maximum lifting capacity of not less than 380 kilograms Floor clearance of not less than inches With integrated charger Weight of not more than 195 kilograms Complies with the following standards: <ul style="list-style-type: none">• BS7177:2008• BS2050• ISO 10993/10		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

FOUR (4) UNITS BRAND NEW ELECTROHYDRAULIC OPERATING TABLES

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(Continuation)</i> <u>FOUR (4) UNITS</u> <u>BRAND NEW</u> <u>ELECTROHYDRAULIC</u> <u>OPERATING TABLES</u>	<ul style="list-style-type: none">• ISO 10993/11• ISO 9001, ISO 13485, ISO 14001:2004• EN 60601-1, EN 60601-2-46, EN 60601-1-2• EC Directive 93/42/EEC Class I• CSA and CE Other terms & conditions of acceptability: 1. The equipment must be a BRAND NEW unit and under warranty including spare parts for at least 2 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts in compliance) with SPMC ISO and CQI Standards. 2. All machines must be manufactured by a known and reputable company with Certificate of Good Manufacturing Practice (GMP), TUV or ISO or its equivalent for equipment only. 3. The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of a Biomedical Technician, End-user and the Company Engineer.		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

FOUR (4) UNITS BRAND NEW ELECTROHYDRAULIC OPERATING TABLES

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>FOUR (4) UNITS</u> <u>BRAND NEW</u> <u>ELECTROHYDRAULIC</u> <u>OPERATING TABLES</u>	<p>4. In line with the ISO standard requirement, the engineer must be capable of conducting annual calibration and the issuance of certificate for the said calibration based on the following data:</p> <ul style="list-style-type: none">a. Reference standardsb. National Institute of Standards and Technology traceabilityc. Reference valuesd. Validity periode. Uncertainty of measurement <p>5. Supplier must comply with the following conditions:</p> <ul style="list-style-type: none">a. Must either be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines.b. Submit Certificate of Training of the Company Engineer / Technical Personnel and Product Specialist issued by the Principal or Manufacturer.c. Provide company response within 24 – 48 hours in case of technical problems or equipment breakdown.		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

FOUR (4) UNITS BRAND NEW ELECTROHYDRAULIC OPERATING TABLES

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><i>(continuation)</i></p> <p><u>FOUR (4) UNITS</u> <u>BRAND NEW</u> <u>ELECTROHYDRAULIC</u> <u>OPERATING TABLES</u></p>	<p>d. In case of machine downtime, the supplier is given three (3) days for remedial action (repair and or replacement of spare parts). On the 4th day, if the unit is not operational, a back-up machine shall be provided (within the warranty period).</p> <p>6. The Supplier must submit the following:</p> <p style="padding-left: 20px;">a. Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment.</p> <p style="padding-left: 20px;">b. Notarized Certificate of availability of spare parts for the next ten (10) years.</p> <p>7. The winning bidder shall provide an Automatic Voltage Regulator suitable for the equipment.</p> <p>8. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end-users and biomedical technician without additional cost to the procuring entity. Supplier must submit a comprehensive training module as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist.</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

FOUR (4) UNITS BRAND NEW ELECTROHYDRAULIC OPERATING TABLES

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity’s Specification</i>	<i>Bidder’s Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>FOUR (4) UNITS</u> <u>BRAND NEW</u> <u>ELECTROHYDRAULIC</u> <u>OPERATING TABLES</u>	9. Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier. 10. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the General Conditions of the Contract (GCC).		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

Name and Signature of Authorized Representative