



Republic of the Philippines
Department of Health
Regional Office No. XI
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

(Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a) (ii) and/or **GCC** Clause 2.1(a) (ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT</u>	<p>PROCUREMENT OF ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT</p> <p>One (1) set Walking Assist Rehabilitation Treadmill With LCD display for time, distance calories, speed, scan LCD Display size of not less than 4 inches With back light 2.0 horsepower motor Speed range of 0.2-8KM/H With six (6) cushion pads With front and rear rollers measuring 70x585mm / 60x530mm Running surface of not less than 510x 1270mm Running deck of not less than 614x1175x25mm</p> <p>One (1) set Interferrential Current with trolley Capable of the following therapies: Biofeedback (EMG and pressure) Combined therapy (stimulation and feedback) Electrotherapy (2 and 4 poles, 2 channels completely independent) Simultaneous therapy (2 different indications treated simultaneously by using electro)</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(Continuation)</i> <u>ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT</u>	With twenty (20) rubber electrodes Safety Class I, ClassIIa, CE compliant One (1) set Biofeedback Controlled Airflow Cryotherapy Mobile cold air device with an adjustable airflow No direct contact with the skin Biofeedback system, a constant treatment temperature Hands free Guided therapy system Integrated indication list, objective list, diagnostics Double-walled, flexible hose with various different attachable nozzles with 1.60 meters in length, rotatable head up to 360 degrees User friendly and clear LCD screen Closed cooling circuit R410A coolant -32 degrees Celsius inside temperature Regulated via biofeedback treatment temperature Maximum airflow of 1000 liters per minute		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(Continuation)</i> <u>ONE (1) LOT</u> <u>BRAND NEW</u> <u>REHABILITATION</u> <u>EQUIPMENT</u>	One (1) set Tilt table Full 90 degrees tilt Standard battery back-up function Adjustable angle and removable work table Adjustable angle split foot board Lowers to wheelchair height Large braking castors Tilt angle indicator 225 kg weight limit Electric height / tilt controls via two function remote With three (3) fixation belts (thoracic, pelvic and knee) One (1) set Multi use Exerciser N-K table Shoulder wheel Overhead pulley Rotary wrist Wrist Roll Finger ladder		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT</u>	One (1) set Knee Continuous Passive Motion Four (4) sets Electrical Stimulator, Transcutaneous Electrical Nerve Stimulator, Functional Electrical Stimulator 4 channels With trolley Electronic circuit, micro controlled Operation modes: TENS: Normal, Burst, VIF, VLP, VLF Pulse frequency of 1 Hz to 200 Hz FES: synchronous, sequential, alternate and reciprocate With twenty (2) rubber electrodes 8x6 cm With 4 exit cables With power cable One (1) set Therapeutic Ultrasound With trolley With multi frequency head, 4cm2 Detachable head		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT</u>	<p>Pulsed and continuous Acoustic and visual contact control</p> <p>One (1) set Cervical and Lumbar traction Static, intermittent and cycling traction User-defined hold, rest and treatment times Ten (10) user-defined protocol slots to store customized settings Digital monochromatic touch screen interface that pivots 270 degrees Easily store treatment parameters, pain scales, pain maps on patient data cards Cervical device</p> <p>One (1) set traction table Two section Friction free gliding lumbar section Traction unit pedestal 181 kilogram lifting capacity Adjustable height between 19 to 18 inches Standard hand control</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>ONE (1) LOT</u> <u>BRAND NEW</u> <u>REHABILITATION</u> <u>EQUIPMENT</u>	One (1) set BAPS Board Multi axial platform system One (1) set powerweb multi resistance Yellow-green Red-blue Ten (10) pieces standard hot moist pack Ten (10) pieces cervical hot moist pack Ten (10) pieces lumbar hot moist pack One (1) set wooble board One (1) set rocker board Two (2) pairs 0.5 lb ankle weights – metal pellets Two (2) pairs 1 lb ankle weights – metal pellets Two (2) pairs 1.5 lbs ankle weights – metal pellets Two (2) pairs 2 lbs ankle weights – metal pellets Two (2) pairs 2.5 lbs ankle weights – metal pellets		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT</u>	Other terms & conditions of acceptability: 1. The equipment must be a BRAND NEW unit and under warranty including spare parts for at least 2 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts in compliance) with SPMC ISO and CQI Standards. 2. All machines must be manufactured by a known and reputable company with Certificate of Good Manufacturing Practice (GMP), TUV or ISO or its equivalent for equipment only. 3. The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of a Biomedical Technician, End-user and the Company Engineer. 4. In line with the ISO standard requirement, the engineer must be capable of conducting annual calibration and the issuance of certificate for the said calibration based on the following data:		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>ONE (1) LOT</u> <u>BRAND NEW</u> <u>REHABILITATION</u> <u>EQUIPMENT</u>	<ul style="list-style-type: none">a. Reference standardsb. National Institute of Standards and Technology traceabilityc. Reference valuesd. Validity periode. Uncertainty of measurement <p>5. Supplier must comply with the following conditions:</p> <ul style="list-style-type: none">a. Must be the exclusive distributor of the principal company of the equipment and the necessary consumables in the Philippines.b. Submit Certificate of Training of the Company Engineer / Technical Personnel and Product Specialist issued by the Principal or Manufacturer.c. Provide company response within 24 – 48 hours in case of technical problems or equipment breakdown.d. In case of machine downtime, the supplier is given three (3) days for remedial action (repair and or replacement of spare parts). On the 4th day, if the unit is not operational, a back-up machine shall be provided (within the warranty period).		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT</u>	<p>6. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end users and biomedical technician without additional cost to the procuring entity. Supplier must submit a comprehensive training module as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist.</p> <p>7. The Supplier must submit the following:</p> <ul style="list-style-type: none">a. Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment.b. Notarized Certificate of availability of spare parts for the next ten (10) years.c. Notarized Certificate of availability of consumables for the next ten (10) years. <p>8. Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier.</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<i>(continuation)</i> <u>ONE (1) LOT BRAND NEW REHABILITATION EQUIPMENT</u>	8. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the General Conditions of the Contract.		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

Name and Signature of Authorized Representative