



Republic of the Philippines
Department of Health
Regional Office No. XI
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATIONS

EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM

(Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a)(ii) and/or **GCC** Clause 2.1(a)(ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
PROCUREMENT OF EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM	<p>EIGHT (8) UNITS WATER TREATMENT AND FILTRATION SYSTEM</p> <p>Can accommodate not less than 750,000 gallons With backwashing features and bypass system NSF certified KDF 55 filtration media With special ultra blend plus With high density garnet bed With breakthrough magnetic field technology</p> <p>EIGHT (8) UNITS AIR PURIFICATION SYSTEM With 10 stage of media air filtration:</p> <ul style="list-style-type: none"> ➤ Polyweb electrostatic filter ➤ Woven polyester ➤ Puratech pre-filter ➤ Micro-grip tackified filter ➤ HEPA type filter ➤ Antimicrobial system ➤ High capacity activated carbon filter ➤ Ultra purification GWM-X filter ➤ Air freshening Ionization stage ➤ Deodorizing & germ elimination ozonator ➤ With remote control 		

Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM

Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p>PROCUREMENT OF EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM <i>(continuation)</i></p>	<p><i>Other terms and conditions for acceptability.</i></p> <ol style="list-style-type: none"> 1. The equipment must be a <i>brand new</i> unit and under warranty including spare parts for at least 2 years. 2. All machines must be manufactured by a known and reputable company with Certificate of Good Manufacturing Practice (GMP), TUV or ISO or its equivalent for equipment only. 3. The supplier must bind itself to service the equipment and maintain the machine on a regular or scheduled basis (at least once in 3 month) in compliance with SPMC ISO and CQI. 4. The equipment/machine must be able to comply with its Installation and Operational Qualification which would be conducted in the presence of an Biomedical Technician with the End-user and the Company Engineer. 5. In line with ISO standard requirement the engineer must be capable of conducting annual calibration certification and provisions of the following data <ol style="list-style-type: none"> a. Reference Standards b. National Institute of Standards and technology (NIST) traceability 		

Signature of Authorized Representative

TECHNICAL SPECIFICATIONS
EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM
Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
PROCUREMENT OF EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM <i>(continuation)</i>	<p><i>Other terms and conditions for acceptability.</i></p> <ul style="list-style-type: none"> c. Reference values d. Validity Period e. Uncertainty of Measurement <p>6. The supplier must comply with the following conditions;</p> <ul style="list-style-type: none"> a. must be either be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines. b. must bind itself to service the equipment within the warranty period of two (2) years and maintain the machine/equipment in the same good condition (e.g. Conduct quarterly preventive maintenance) at its own expense including labor and cost of spare parts. c. must have a Service Engineer or Technical personnel based in Davao City to do preventive maintenance and answer client calls in case of technical problems or concerns. d. Certificate of Training of the Company Engineer / Technical Personnel issued by the Principal or Manufacturer. e. provide company response within 24-48hrs should the existing equipment breaks down. 		

Signature of Authorized Representative

TECHNICAL SPECIFICATIONS
EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM
Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non-responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p>PROCUREMENT OF EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM <i>(continuation)</i></p>	<p><u>Other terms and conditions for acceptability.</u></p> <p style="padding-left: 40px;">f. provide a substitute trained technical personnel/specialist in cases when the service engineer is out of town or is not available.</p> <p style="padding-left: 40px;">g. In case of machine downtime, the supplier is given 2 days to do remedial action (repair). On the 3rd day, if the unit is not operational, a penalty shall be imposed equivalent to the average daily income of the hospital derive from the affected machine.</p> <p>7. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end users and biomedical technician without additional cost to the procuring entity.</p> <p>8. Supplier shall provide three (3) operating manuals and three (3) service manuals in English language.</p>		

 Signature of Authorized Representative

TECHNICAL SPECIFICATIONS
EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM
Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
PROCUREMENT OF EIGHT (8) UNITS BRAND NEW WATER TREATMENT FILTRATION & EIGHT (8) UNITS AIR PURIFICATION SYSTEM <i>(continuation)</i>	<u>Other terms and conditions for acceptability.</u> 9. Supplier will be responsible for the planning, installation of the water treatment filtration system including electrical, plumbing and civil works without additional cost to the procuring entity. 10. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the GCC.		

The bidder may submit offer which provides for superior specification and/or better terms and conditions to the government. However, these shall not be given any bonus, credit or premium in the bid evaluation.

 Signature of Authorized Representative