



Republic of the Philippines
Department of Health
Regional Office No. XI
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF TWO (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR

(Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a) (ii) and/or **GCC** Clause 2.1(a) (ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF TWO (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF TWO (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR</u>	<ul style="list-style-type: none">• <u>TWO UNITS ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR</u> <p>ANESTHESIA MACHINE SPECIFICATIONS:</p> <p>PHYSICAL:</p> <ol style="list-style-type: none">1. 3- Gas Version O2/N2O/Air2. 7.5" TFT display screen with trim knob manipulation3. Folding side shelf for large working space and task light4. 3 large drawers with 4 caster wheels and brakes5. Dimensions: Height: 145 – 148 cm Width 87 – 95 cm Height 67 – 70 cm <p>VENTILATOR:</p> <ol style="list-style-type: none">1. Vent Modes: Volume, Pressure, SIMV-VC, SIMV-PC and Pressure Support with Apnea Backup2. Tidal Volume Range: 20-1500 ml3. Flow Trigger: 0.2 to 10 L/min4. Electronic PEEP: Off to 25 cmH2O5. Customizable Alram Settings		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR</u>	<ol style="list-style-type: none"> 6. 2 flow sensors and galvanic fuel cell oxygen sensor 7. Capable of Spirometry 8. With not less than 90 minutes battery backup <p>VAPORIZER:</p> <ol style="list-style-type: none"> 1. 2 Manifold compatible with 3 years calibration-free Selectatec Vaporizers 2. Tool-free installation and manifold interlocks <p>PNEUMATIC AND BREATHING CIRCUIT SPECIFICATIONS</p> <ol style="list-style-type: none"> 1. Proportionate decrease of N2O with reduction of O2 pressure 2. Mechanical gear hypoxic guard system providing a nominal minimum 25% concentration of Oxygen in O2/N2O mixture 3. O2 Flush range: 25-75 L/min 4. I:E Ratio range: 2:1 to 1:18 5. 2- 3 gas delivery system 6. Circuit Volume of not more than 2.6L Vent mode (absorber included) 7. Absorbent Capacity of not more than 1200 ml 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR</u>	<ul style="list-style-type: none">8. APL Valve with tactile knob indication at 30cm H2O and above9. All materials in contact with exhaled patient gases are autoclavable, except disposable flow sensors and O2 cell10. All materials in contact with patient gas are free of natural rubber latex11. No change Bellows for neonate to adult patients12. No tools needed for assembly and disassembly of Bellows system		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR</u>	PATIENT MONITOR SPECIFICATIONS: TECHNICAL: <ol style="list-style-type: none">1. 10-15" capacitive touchscreen displays with knob manipulation and soft keys2. Can display up to 8 waveform traces up to 14 with overlays and insets and 8 digital values display3. Central audible and visual alarm display4. All parameter trends selectable time scales from 20 min to 72hour trend display5. Maximum 6 kg weight with integrated transport handle6. Battery life of 2 hours with option for 1 battery pack slot7. Capable of storing ≥ 400 snapshots and ≥ 999 events8. With at least 8 configurable profiles and 6 configurable pages per profile for parameter setup and monitoring9. With integrated carrying handle10. At least 2 module slots for additional special parameters11. Option for centralized viewing and remote alarm management with bed to bed viewing and automatic view on alarm functionality		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR</u>	PARAMETERS: <ol style="list-style-type: none">1. ECG: 3, 5 or 12 lead ecg configuration with Pacemaker Detection, can save ≥ 15 ECG reports2. SP02: Waterproof and shock proof finger sensor3. NIBP: Oscillometric with step deflation technique; Customizable NIBP Series up to 4 series4. Invasive Pressure: 2-4 ports5. Temperature: Skin and Core temperature monitoring6. Capable of Cardiac Output monitoring7. Heart Rate: 30 – 300 bpm8. Impedance respiration monitoring is integrated9. Respiratory Module with accessories integrated in the system10. Neuromuscular Transmission Module with accessories integrated in the system		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR</u></p> <p><i>(continuation)</i></p>	<p>Other terms & conditions of acceptability:</p> <ol style="list-style-type: none"> 1. The equipment must be BRAND-NEW and under warranty including spare parts for at least 3 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts). 2. All machines must be manufactured by a known and reputable company compliant with ISO or its equivalent for equipment only. 3. The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of SPMC Biomedical Technician, End-user and the Company Engineer. 4. In line with the ISO standard requirement, the supplier's engineer or technical personnel must be capable of conducting annual calibration and the issuance of certificate for the said calibration based on the following data: <ol style="list-style-type: none"> a. Reference standards b. National Institute of Standards and Technology traceability c. Reference values d. Validity period e. Uncertainty of measurement 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR</u></p> <p><i>(continuation)</i></p>	<p>5. Supplier must comply with the following conditions:</p> <ul style="list-style-type: none"> a. Must be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines. b. Submit Certificate of Training of the Company Engineer / Technical Personnel and Product Specialist issued by the Principal or Manufacturer. c. Provide company response within 24 – 48 hours in case of technical problems or equipment breakdown. a. d. In case of machine downtime, the supplier is given three (3) days for remedial action (repair and or replacement of spare parts). On the 4th day, if the unit is not operational, a back-up machine shall be provided within the warranty period. The supplier will be given fourteen (14) days to comply in providing back-up machine. A penalty will be imposed after the fourteenth day amounting to the average daily income of the machine. This is to cover for the income loss. <p>6. The Supplier must submit the following:</p> <ul style="list-style-type: none"> b. Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment. c. Notarized Certificate of availability of spare parts for the next ten (10) years. 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (2) UNITS BRAND NEW ANESTHESIA VENTILATOR MACHINE WITH PATIENT MONITOR</u></p> <p><i>(continuation)</i></p>	<p>7. Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier.</p> <p>8. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end-users and biomedical technician without additional cost to the procuring entity. Supplier must submit a comprehensive training module as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist. The module should indicate the following:</p> <ul style="list-style-type: none"> - The objectives of the training - The methods or activities that shall be done during the training - The training outline that indicates the concepts and/or skills to be discussed/ taught. - The target participants of the training - The activity schedule indicating the no. of days or hours <p>9. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the General Conditions of the Contract (GCC).</p>		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

Name and Signature of Authorized Representative