



Republic of the Philippines
Department of Health
Regional Office No. XI
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT **BRAND NEW ENT EQUIPMENT PACKAGE** (Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a) (ii) and/or **GCC** Clause 2.1(a) (ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT **BRAND NEW ENT EQUIPMENT PACKAGE**

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY,</u> <u>DELIVERY AND</u> <u>INSTALLATION OF</u> <u>ONE (1) LOT</u> <u>BRAND NEW ENT</u> <u>EQUIPMENT</u> <u>PACKAGE</u>	<ol style="list-style-type: none">1. One (1) pc ROBERTS-JESSBERG Esophagoscope, oval, size 12x16mm, length 50-55cm2. One (1) pc ROBERTS-JESSBERG Esophagoscope, oval, size 10x14mm, length 50-55cm3. One (1) pc Fiber optic light carrier, for use with esophagoscopes, length 50-55cm4. One (1) pc Suction Tube, for use with esophagoscopes, outer diameter 2mm, length 50-55cm5. One (1) pc Metal Handle, for esophagoscopes use6. One (1) pc HOPKINS straight forward telescope 0 degrees, diameter 5-5.5mm, length 53-55cm, autoclavable, fiber optic light transmission incorporated7. One (1) pc Forceps, alligator, for hard foreign bodies, single action jaws, sheath diameter 2-2.5mm, working length 55-60mm8. One (1) pc Forceps, alligator, for peanuts and soft foreign bodies, single action jaws, sheath diameter 2-2.5mm, working length 55-60cm9. One (1) pc Forceps, alligator, universal for biopsy and foreign body removal, single action jaws, sheath diameter 2-2.5mm, working length 55-60cm		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT **BRAND NEW ENT EQUIPMENT PACKAGE**

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY,</u> <u>DELIVERY AND</u> <u>INSTALLATION OF</u> <u>ONE (1) LOT</u> <u>BRAND NEW ENT</u> <u>EQUIPMENT</u> <u>PACKAGE</u> <i>(continuation)</i>	Other terms & conditions of acceptability: 1. The equipment must be BRAND-NEW and under warranty including spare parts for at least 2 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts). 2. All machines must be manufactured by a known and reputable company compliant with ISO or its equivalent for equipment only. 3. The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of SPMC Biomedical Technician, End-user and the Company Engineer. 4. In line with the ISO standard requirement, the supplier's engineer or technical personnel must be capable of conducting annual calibration and the issuance of certificate for the said calibration based on the following data: <div style="margin-left: 40px;"><div>a. Reference standards</div><div>b. National Institute of Standards and Technology traceability</div><div>c. Reference values</div><div>d. Validity period</div><div>e. Uncertainty of measurement</div></div>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT **BRAND NEW ENT EQUIPMENT PACKAGE**

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT EQUIPMENT PACKAGE</u> <i>(continuation)</i>	<p>5. Supplier must comply with the following conditions:</p> <ul style="list-style-type: none"> a. Must be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines. b. Submit Certificate of Training of the Company Engineer / Technical Personnel and Product Specialist issued by the Principal or Manufacturer. c. Provide company response within 24 – 48 hours in case of technical problems or equipment breakdown. d. In case of machine downtime, the supplier is given three (3) days for remedial action (repair and or replacement of spare parts). On the 4th day, if the unit is not operational, a back-up machine shall be provided within the warranty period. The supplier will be given 14 days to comply in providing back up machine. A penalty will be imposed after the 14th day amounting to the average daily income of the machine. This is to cover for the income loss. <p>6. The Supplier must submit the following:</p> <ul style="list-style-type: none"> a. Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment. b. Notarized Certificate of availability of spare parts for the next ten (10) years. <p>7. Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier.</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT EQUIPMENT PACKAGE

Name of Project


The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT EQUIPMENT PACKAGE</u></p> <p><i>(continuation)</i></p>	<p>8. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end-users and biomedical technician without additional cost to the procuring entity. Supplier must submit a comprehensive training module as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist. The module should indicate the following:</p> <ul style="list-style-type: none"> - The objectives of the training - The methods or activities that shall be done during the training - The training outline that indicates the concepts and/or skills to be discussed/ taught. - The target participants of the training - The activity schedule indicating the no. of days or hours <p>9. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the General Conditions of the Contract (GCC).</p>		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

Name and Signature of Authorized Representative

11.2.3 Definition of Axis 1–Axis 6



11.2.4 System dimensions

