



Republic of the Philippines
Department of Health
Regional Office No. XI
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

(Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a) (ii) and/or **GCC** Clause 2.1(a) (ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u>	<ol style="list-style-type: none"> 1. One (1) pc Cotton Carrier, for endolaryngeal microsurgery, straight, working length 18 - 20cm 2. One (1) pc Cotton Carrier, straight, working length 25 -28cm. 3. One (1) pc Cotton Carrier, for endolaryngeal microsurgery, straight, working length 21 -24cm 4. One (1) pc FEHLAND Needle Holder, small model, for phonosurgery and endolaryngeal microsurgery, with ratchet, sheath diameter 3 mm, jaws size 2 x 4 mm, angled 90° to the right, with cleaning connector, working length 20-23 cm 5. One (1) pc FEHLAND Needle Holder, small model, for phonosurgery and endolaryngeal microsurgery, with ratchet, sheath diameter 3 mm, jaws size 2 x 4 mm, angled 90° to the left, with cleaning connector, working length 20-23 cm 6. One (1) pc LINDHOLM Forceps, for atraumatic retraction of true vocal cords and false vocal cords, distal end with blunt curved blades, self-retaining, with ratchet and cleaning connector, working length 24 -26cm 7. One (1) pc Sickle Knife curved, delicate, 3 to 1 mm tapered shaft, total length 23 cm 8. One (1) pc Sickle Knife straight cutting, delicate, 3 to 1 mm tapered shaft, total length 23 cm 9. One (1) pc Round Knife, vertical cutting, delicate, 3 to 1 mm tapered shaft, total length 23 cm 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>10. One (1) pc Needle straight, sharp, delicate, 3 to 1 mm tapered shaft, total length 23 cm</p> <p>11. One (1) pc Hook 90° 1 mm, sharp, delicate, 3 to 1 mm tapered shaft, total length 23 cm</p> <p>12. One (1) pc Hook 90° 2 mm, sharp, delicate, 3 to 1 mm tapered shaft, total length 23 cm</p> <p>13. One (1) pc Dissector straight, ball end, delicate, 3 to 0.6 mm tapered shaft, diameter 1.2 - 1.3 mm, total length 23 cm</p> <p>14. One (1) pc Suction Elevator, angular, with grip plate and cut-off hole, LUER, distal elevator blade curved to left delicate, tapered sheath, working length 21-24 cm, total length 23 cm</p> <p>15. One (1) pc Suction Elevator, angular, with grip plate and cut-off hole, LUER, distal elevator blade curved to right, delicate, tapered sheath, working length 21-24 cm, total length 23 cm</p> <p>16. One (1) pc Elevator slightly curved, blunt, delicate, 3 to 1 mm tapered shaft, total length 23 cm</p> <p>17. One (1) pc Universal Handle, forceps shape, with 2 handle rings, for use with 3mm-instruments for phonosurgery and endolaryngeal microsurgery, length 9 cm</p> <p>18. One (1) pc Elevator slightly curved, sharp tip, delicate, 3 to 1 mm tapered shaft, total length 23 cm</p> <p>19. One (1) pc Elevator angled 15°, sharp, delicate, 3 to 1 mm tapered shaft, total length 23 cm</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>20. One (1) pc Curette oval spoon, straight, delicate, 3 to 1 mm tapered shaft, total length 23 cm</p> <p>21. One (1) pc Curette, oval spoon, angled 35°, delicate, 3 mm to 1 mm, tapered shaft, total length 23 cm</p> <p>22. One (1) pc Protector, to protect tissue against unintentional laser irradiation, curved upwards, oval, diameter 6-8 mm, special matt finish, with suction channel to remove laser vapors, working length 23-25 cm</p> <p>23. One (1) pc Metal Handle, without ratchet</p> <p>24. One (1) pc KLEINSASSER Forceps, straight, with 2 mm cupped jaws, with cleaning connector, working length 23 -25cm</p> <p>25. One (1) pc KLEINSASSER Forceps, curved upwards, with 2 mm cupped jaws, with cleaning connector, working length 23 -25cm</p> <p>26. One (1) pc KLEINSASSER Forceps, curved to right, with 2 mm cupped jaws, with cleaning connector, working length 23 -25cm</p> <p>27. One (1) pc KLEINSASSER Forceps, curved to left, with 2 mm cupped jaws, with cleaning connector, working length 23 -25cm</p> <p>28. One (1) pc KLEINSASSER Grasping Forceps, without ratchet, serrated, straight, with cleaning connector, straight, working length 23 -25cm</p> <p>29. One (1) pc KLEINSASSER Grasping Forceps, without ratchet, serrated, curved to right, with cleaning connector, working length 23 -25cm</p> <p>30. One (1) pc KLEINSASSER Grasping Forceps, without ratchet, serrated, curved to left, with cleaning connector, working length 23 -25cm</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>31. One (1) pc KLEINSASSER Grasping Forceps, without ratchet, serrated, curved upwards, with cleaning connector, working length 23 -25cm</p> <p>32. One (1) pc KLEINSASSER Artery Forceps, with ratchet, with cleaning connector, working length 23 -25cm</p> <p>33. One (1) pc Forceps, spoon-shaped, oval, width 4 mm, fenestrated, for biopsy, double action jaws, sheath diameter 2.3 - 2.5 mm, with cleaning connector, working length 23 -25cm</p> <p>34. One (1) pc Spoon Forceps, with cupped jaws, diameter 5 mm, robust model, double action jaws, sheath diameter 2.5 mm, with cleaning connector, working length 23 -25cm</p> <p>35. One (1) pc Instrument Rack, with two (2) irrigation blocks, for cleaning, sterilization and storage of up to 30 instrument attachments and 6 handles, incl. variable bars with silicone holder and direct connection of the instrument attachments via a multi-port bridge to the washer-disinfector and Wire Tray, external dimensions (w x d x h): approximately 480 x 250 x 160 mm</p> <p>36. One (1) pc KLEINSASSER Suction Tube, ball end, outer diameter 3 mm, working length 18-20 cm</p> <p>37. One (1) pc Protector, to protect tissue against unintentional LASER irradiation, curved upwards, round, diameter 5 mm, special matte finish, with suction channel to remove LASER vapors, working length 23 -25cm</p> <p>38. One (1) pc Coagulation Suction Tube, with ergonomic handle, insulated, with connection for unipolar coagulation, with cleaning stylet, O.D. 2 mm, working length 23 -25cm</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>39. One (1) pc Coagulation Suction Tube, with ergonomic handle, insulated, with connection for unipolar coagulation, with cleaning stylet O.D. 2.5 mm, working length 23-25 cm</p> <p>40. One (1) pc Suction Tube, with ergonomic handle, with distal elevator 18 x 5 mm, for protecting and retracting tissue, special matt finish, with Cleaning Stylet, outer diameter 2.5 - 3 mm, working length 23 -25cm</p> <p>41. One (1) pc Suction Tube, ball end, O.D. 2 -2.5 mm working length 18 -20cm.</p> <p>42. One (1) pc Injection Cannula, short tip, for single use, LUER-Lock, tip outer diameter 1 mm, inner diameter 0.7 mm, working length 18 cm, package of 3</p> <p>43. One (1) pc Teeth Protector, metal, large</p> <p>44. One (1) pc Teeth Protector, metal, medium</p> <p>45. One (1) pc Teeth Protector, metal, small</p> <p>46. One (1) pc Teeth Protector, metal, extra small</p> <p>47. One (1) pc HOPKINS Straight Forward Telescope 15° diameter 4 mm, length 17 cm, autoclavable, 45° angled eyepiece, fiber optic light transmission incorporated, color code: green</p> <p>48. One (1) pc KANTOR-BERCI Video-Laryngoscope, model IV, for adolescents and adults, length 17 cm, enlarged proximal opening, curved distal tip especially designed to elevate the epiglottis and for optimal placement of the anterior commissure, flattened surface of the blade prevent damage of teeth, telescope guide tube integrated inside the lumen, with adaptor for HOPKINS telescope, for use with proximal light clip and laryngoscope holder</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>49. One (1) pc KLEINSASSER Laryngoscope, medium, modified by RUDERT for CO2-LASER surgery, for adults, with lateral outer channels for suction tube to remove vapor and/or for fiber optic light carrier, length 18 cm, for anterior commissure</p> <p>50. One (1) pc KLEINSASSER Spoon Forceps, 2 mm, straight, sheath conically reinforced from distal to proximal end, distal end curved upwards 10°, with cleaning connector, working length 23 -25cm</p> <p>51. One (1) pc KLEINSASSER Spoon Forceps, 2 mm, curved upwards 45°, sheath conically reinforced from distal to proximal end, distal end curved upwards 10°, with cleaning connector, working length 23-25 cm</p> <p>52. One (1) pc KLEINSASSER Suction Tube ball end, curved upwards, outer diameter 2 -2.3mm, working length 23- 25cm</p> <p>53. One (1) pc KLEINSASSER Suction Tube ball end, curved upwards, outer diameter 3-3.3 mm, working length 23 -25cm</p> <p>54. One (1) pc Clip for proximal illumination</p> <p>55. One (1) pc Injection Cannula for positive pressure assisted ventilation, working length 8.5 -10cm, for use with Laryngoscopes</p> <p>56. One (1) pc Suction Tube to remove vapor for LASER treatment, diameter 3 - 3.3mm, working length 12.5 -14.5cm, for use with Laryngoscopes</p> <p>57. One (1) pc Fiber Optic Light Carrier, for distal illumination, working length 14- 16 cm, for use with Laryngoscopes and Diverticulosopes</p> <p>58. One (1) pc LINDHOLM Operating Laryngoscope, large, for adults, inner proximal end 39 x 24 mm, distal end width 18-20 mm, length 15 -17cm</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>59. One (1) pc LINDHOLM Operating Laryngoscope length 11-13 cm, pediatric size, distal end 16-18 mm wide.</p> <p>60. One (1) pc BENJAMIN-LINDHOLM Laryngoscope, small pediatric size, length 9-9.5 cm, inner diameter proximal 26 mm x 16 mm, distal 14 mm x 11 mm</p> <p>61. One (1) pc PARSONS Operating Laryngoscope, size 6, for adults, large, flat elevated tip for better positioning in the tongue ground and anterior commissure, wide proximal opening for binocular viewing, right-hand side slotted for introduction of Broncho-esophagoscopes and operating instruments, left-hand side with insufflation channel, for use with clip for proximal illumination, length 18-20 cm</p> <p>62. One (1) pc PARSONS Operating Laryngoscope, size 4, children up to age 12, medium, flat elevated tip for better positioning in the tongue ground and anterior commissure, wide proximal opening for binocular viewing, right-hand side slotted for introduction of Broncho-esophagoscopes and operating instruments, left-hand side with insufflation channel, for use with clip for proximal illumination, length 14.5 -16.5cm</p> <p>63. One (1) pc PARSONS Laryngoscope, size 3, for toddlers, flat distal end for anterior commissure, wide proximal opening enable binocular viewing, right-hand side slotted for introduction of Broncho-oesophagoscopes and operating instruments, left-hand side with insufflation channel for oxygen supply, anesthetic gas or jet ventilation, for use with Prismatic Light Deflector and Laryngoscope Holder, length 11 -13cm</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>64. One (1) pc PARSONS Laryngoscope, size 2, for neonates, flat distal end for anterior commissure, wide proximal opening enable binocular viewing, right-hand side slotted for introduction of Broncho-oesophagoscopes and operating instruments, left-hand side with insufflation channel for oxygen supply, anesthetic gas or jet ventilation, for use with Prismatic Light Deflector and Laryngoscope Holder, length 9-11 cm</p> <p>65. One (1) pc PARSONS Laryngoscope, size 1, for premature infants, flat distal end for anterior commissure, wide proximal opening enable binocular viewing, right-hand side slotted for introduction of Broncho-oesophagoscopes and operating instruments, left-hand side with insufflation channel for oxygen supply, anesthetic gas or jet ventilation, for use with Prismatic Light Deflector and Laryngoscope Holder, length 8-10cm</p> <p>66. One (1) pc Prismatic Light Deflector autoclavable, with connection to fiber optic light cable</p> <p>67. One (1) pc BENJAMIN-PARSONS Laryngoscope Holder and Chest Support, for use with pediatric laryngoscopes with proximal illumination, consisting of: Laryngoscope Holder GOETTINGEN model, with adjustment wheel</p> <p>68. One (1) pc BENJAMIN-PARSONS Support Rod, moveable, with metal ring ø 12 cm and 2 lateral screws, length 34-36 cm</p> <p>69. One (1) pc Laryngeal Probe, with handle, angled shaft, straight tip, working length 23 -25cm</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>70. One (1) pc Biopsy Forceps flexible, double action jaws, oval, diameter 1-1.5 mm, length 60-65 cm</p> <p>71. One (1) pc Grasping Forceps flexible, double action jaws, diameter 1 -1.5mm, length 60 -65cm</p> <p>72. One (1) pc HOPKINS Straight Forward Telescope 0°, enlarged view, diameter 4.5-5.5 mm, length 24 -26cm, autoclavable. Fiber optic light transmission incorporated. Color code: green</p> <p>73. One (1) pc HOPKINS Forward Oblique Telescope 30°, enlarged view, diameter 4.5-5.5 mm, length 24 -26cm, autoclavable. Fiber optic light transmission incorporated. Color code: red</p> <p>74. One (1) pc HOPKINS Straight Forward Telescope 0°, enlarged view, diameter 5 mm, length 29 cm, autoclavable, fiber optic light transmission incorporated, color code: green</p> <p>75. One (1) pc KLEINSASSER Needle Holder, delicate, straight, serrated jaws, size 1.8 x 3.5 mm, sheath conically reinforced from distal to proximal end, with ratchet, with cleaning connector, working length 23 cm</p> <p>76. One (1) pc Protector, to protect tissue against unintentional LASER irradiation, curved upwards, round, diameter 5 mm, special matt finish, with suction channel to remove LASER vapors, working length 23 cm</p> <p>77. One (1) pc Protector, to protect tissue against unintentional LASER irradiation, curved upwards, oval, diameter 7 mm, special matt finish, with suction channel to remove LASER vapors, working length 23 cm</p> <p>78. One (1) pc KLEINSASSER Miniature Scissors, straight, with cleaning connector, working length 23 cm</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>		<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>79. One (1) pc KLEINSASSER Miniature Scissors, 45° upturned, with cleaning connector, working length 23 cm</p> <p>80. One (1) pc KLEINSASSER Miniature Scissors, curved to right, with cleaning connector, working length 23 cm</p> <p>81. One (1) pc KLEINSASSER Miniature Scissors, curved to left, with cleaning connector, working length 23 cm</p> <p>82. One (1) pc Chest Support, LÜBECK model, with gear system for height adjustment and swivel arm with movable plate, for use with Laryngoscope Chest Supports, can be mounted on OR table equipped with standard sliding rail 25 x 10 mm</p> <p>83. One (1) pc Extension GRONINGEN Model for Laryngoscope Holder, angular, with screw for adaption to adult operating laryngoscopes, to enlarge opening angle of the chest support rod for obese patients</p> <p>84. One (1) pc HOPKINS Straight Forward Telescope 0°, enlarged view, diameter 5 mm, length 29 cm, autoclavable, fiber optic light transmission incorporated, color code: green</p> <p>85. One (1) pc Endoscopic video unit for use with one-chip camera heads and video endoscopes, including LED light source similar to Xenon technology, with integrated digital image</p> <ul style="list-style-type: none"> - processing "module 15", LCD TFT monitor with LED backlight, USB/SD memory module, color systems PAL/NTSC, power supply 100-240VAC, 50/60Hz including: USB silicone keyboard, with touchpad, USB flash drive, 32GB TP 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>86. One (1) pc 1-chip camera head, with 2 freely programmable camera head buttons, color system NTSC, with integrated Parfocal-Zoom - focal length = 25-50mm (2x)</p> <p>87. One (1) pc Fiber optic light cable, with straight connector, extremely heat-resistant, with safety lock, diameter 4.8mm, length 250cm</p> <p>88. One (1) pc Laryngoscope holder and chest support, GOTTINGEN model, consisting of: laryngoscope holder, GOTTINGEN model, with adjustment wheel, support rod, movable with metal ring, diameter 9cm, length 34cm</p> <p>89. One (1) pc Prismatic light deflector, autoclavable, with fiber optic light cable, for use with pediatric laryngoscopes</p> <p>90. One (1) pc Extension GRONINGEN model for laryngoscope holder, angular, with screw for adaptation to adult operating laryngoscopes, to enlarge opening angle of the chest support rod for obese patients</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>Other terms & conditions of acceptability:</p> <ol style="list-style-type: none"> 1. The equipment must be BRAND-NEW and under warranty including spare parts for at least 2 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts). 2. All machines must be manufactured by a known and reputable company compliant with ISO or its equivalent for equipment only. 3. The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of SPMC Biomedical Technician, End-user and the Company Engineer. 4. In line with the ISO standard requirement, the supplier's engineer or technical personnel must be capable of conducting annual calibration and the issuance of certificate for the said calibration based on the following data: <ol style="list-style-type: none"> a. Reference standards b. National Institute of Standards and Technology traceability c. Reference values d. Validity period e. Uncertainty of measurement 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>5. Supplier must comply with the following conditions:</p> <ul style="list-style-type: none"> a. Must be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines. b. Submit Certificate of Training of the Company Engineer / Technical Personnel and Product Specialist issued by the Principal or Manufacturer. c. Provide company response within 24 – 48 hours in case of technical problems or equipment breakdown. d. In case of machine downtime, the supplier is given three (3) days for remedial action (repair and or replacement of spare parts). On the 4th day, if the unit is not operational, a back-up machine shall be provided within the warranty period. The supplier will be given 14 days to comply in providing back up machine. A penalty will be imposed after the 14th day amounting to the average daily income of the machine. This is to cover for the income loss. <p>6. The Supplier must submit the following:</p> <ul style="list-style-type: none"> a. Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment. b. Notarized Certificate of availability of spare parts for the next ten (10) years. <p>7. Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier.</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>8. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end-users and biomedical technician without additional cost to the procuring entity. Supplier must submit a comprehensive training module as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist. The module should indicate the following:</p> <ul style="list-style-type: none"> • The objectives of the training • The methods or activities that shall be done during the training • The training outline that indicates the concepts and/or skills to be discussed/ taught. • The target participants of the training • The activity schedule indicating the no. of days or hours <p>9. The supplier must provide training mannequin (neonatal or pediatric) with oral and nasal access for use for laryngoscopes.</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p style="text-align: center;"><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) LOT BRAND NEW ENT INSTRUMENTS FOR AIRWAY CENTER</u></p> <p style="text-align: center;"><i>(continuation)</i></p>	<p>10. The supplier must provide Desktop computer with the following specifications:</p> <ul style="list-style-type: none"> • Processor: Intel Core i5 7th gen or higher • RAM: 8GB DDR4 or higher • HDD: 1TB Seagate Hard Disk, 7.21rpm SATA or higher • Motherboard with built-in Gigabit LAN Sound and Video card with PCI Wireless LAN card • CPU case with PSU and DVD: Writer • LED Monitor, at least 21" • Mouse and Keyboard USB type • Backup Battery UPS: 600VA • AVR • Windows 10 Professional 64bit with media kit or higher • Microsoft Office Home and Business Edition 2016 • Laser jet Printer <p>11. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the General Conditions of the Contract (GCC).</p>		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

Name and Signature of Authorized Representative

