



Republic of the Philippines
Department of Health
Regional Office No. XI
SOUTHERN PHILIPPINES MEDICAL CENTER
J.P. Laurel Avenue, Davao City



TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE

(Name of Project)

Instructions:

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a) (ii) and/or **GCC** Clause 2.1(a) (ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

Prepared by Technical Working Group

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE</u>	<ul style="list-style-type: none">• <u>ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE</u><ol style="list-style-type: none">1. Type: Class I, Type B electrical equipment2. Photographic field: within 0.25 x 0.55 mm3. Photographic capability: Automatic4. Photographic location: Center, Peripheral location (0, 45, 135, 180, 225, 315°)5. Central corneal thickness: Measurable, using non-contact pachymetry6. Flash: LED7. Analytical method: Fully automated, Semi automated, Center Method and Flex-Center Method8. Analysis result: Cell density (mm²), Coefficient of variation, Standard deviation, Percentage of hexagonal cells, Number of calculated cells, Average cell area (µm²)9. Output function: USB 3.0 x 4, LAN, Bluetooth, Wireless LAN10. Integrated monitor: at least 10.5;" Capacitive touch panel with a 180 ° flexible horizontal and vertical movement capability11. Center method: Gold standard center and flex-center method		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE</u> <i>(continuation)</i>	12. Connections: EMR connection/ DICOM capability 13. Power supply: 220-240 VAC, 60 Hz 14. Dimensions: not more than 310 (W) X 460 (D) X 455 (H) mm (with the monitor at the rear-side) 15. Weight: not more than 20kg 16. Accessories a. Printer: Continuous ink b. Uninterrupted Power Supply: at least 150% of total equipment power capacity c. Two (2) units Stools i. Pneumatic assist seat adjustment ii. Five-leg base for extra stability iii. 13-15" wide circular seat cushion d. One (1) unit Steel cabinet, Vertical 3 drawer i. All metal construction, steel ball bearing, gauge – 0.9mm or gauge 20 ii. Dimensions: Width – 45-50cm; Length – 55-60cm; Height – 100-105 cm iii. Accessories: file divider iv. Central locking system e. Provide spare LED Bulb, four (4) pieces		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE</u></p> <p><i>(continuation)</i></p>	<p>Other terms & conditions of acceptability:</p> <ol style="list-style-type: none"> 1. The equipment must be BRAND-NEW and under warranty including spare parts for at least 3 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts). 2. All machines must be manufactured by a known and reputable company compliant with ISO or its equivalent for equipment only. 3. The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of SPMC Biomedical Technician, End-user and the Company Engineer. 4. In line with the ISO standard requirement, the supplier's engineer or technical personnel must be capable of conducting annual calibration and the issuance of certificate for the said calibration based on the following data: <ol style="list-style-type: none"> a. Reference standards b. National Institute of Standards and Technology traceability c. Reference values d. Validity period e. Uncertainty of measurement 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE</u></p> <p><i>(continuation)</i></p>	<p>5. Supplier must comply with the following conditions:</p> <ul style="list-style-type: none"> a. Must be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines. b. Submit Certificate of Training of the Company Engineer / Technical Personnel and Product Specialist issued by the Principal or Manufacturer. c. Provide company response within 24 – 48 hours in case of technical problems or equipment breakdown. d. In case of machine downtime, the supplier is given three (3) days for remedial action (repair and or replacement of spare parts). On the 4th day, if the unit is not operational, a back-up machine shall be provided within the warranty period. The supplier will be given fourteen (14) days to comply in providing back-up machine. A penalty will be imposed after the fourteenth day amounting to the average daily income of the machine. This is to cover for the income loss. <p>6. The Supplier must submit the following:</p> <ul style="list-style-type: none"> a. Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment. b. Notarized Certificate of availability of spare parts for the next ten (10) years. <p>7. Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier.</p>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE</u></p> <p><i>(continuation)</i></p>	<p>8. Supplier will arrange for the furnishing of the OR Reception Counter, Inclusive of the following:</p> <ul style="list-style-type: none"> • One (1) set Desktop Computer <ul style="list-style-type: none"> i. Intel i7 9700k processor ii. Memory: 8gb iii. HDD: 1TB iv. SSD: 240GB v. OS: Windows 10 vi. Monitor: HD 1920x 1080 vii. Wireless keyboard viii. Wireless mouse • Printer: Tank type refillable • Three (3) units Office chairs • Three (3) units Cabinets for filing • Personalized wall mounted Department logo, made in acrylic with internal lighting • Intercom speaker system 		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE</u> <i>(continuation)</i>	9. Supplier will arrange for the furnishing of the Releasing/Receiving Counter to/from OPD, inclusive of the following: <ul style="list-style-type: none">One (1) set Desktop Computer<ul style="list-style-type: none">Intel i7 9700k processorMemory: 8gbHDD: 1TBSSD: 240GBOS: Windows 10Monitor: HD 1920x 1080Wireless keyboardWireless mousePrinter: Tank type refillableThree (3) Office chairsTwo (2) cabinets for filing		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<p><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT BRAND NEW SPECULAR MICROSCOPE</u></p> <p><i>(continuation)</i></p>	<p>10. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end-users and biomedical technician without additional cost to the procuring entity. Supplier must submit a comprehensive training module as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist. The module should indicate the following:</p> <ul style="list-style-type: none"> - The objectives of the training - The methods or activities that shall be done during the training - The training outline that indicates the concepts and/or skills to be discussed/ taught. - The target participants of the training - The activity schedule indicating the no. of days or hours <p>11. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the General Conditions of the Contract (GCC).</p>		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

Name and Signature of Authorized Representative