



Republic of the Philippines  
Department of Health  
Regional Office No. XI  
**SOUTHERN PHILIPPINES MEDICAL CENTER**  
J.P. Laurel Avenue, Davao City



## TECHNICAL SPECIFICATIONS

### **SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES**

(Name of Project)

#### ***Instructions:***

Bidders must state in the column under **Statement of Compliance** the word either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of **ITB** Clause 3.1(a) (ii) and/or **GCC** Clause 2.1(a) (ii).

Bidder may submit offer which provides for superior specifications and /or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation

**Prepared by Technical Working Group**

## TECHNICAL SPECIFICATIONS

### SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<b><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</u></b>	<ul style="list-style-type: none"> <li>• <b><u>ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT COMPATIBILITY</u></b></li> </ul> <p>OPTICS AND ILLUMINATION</p> <ol style="list-style-type: none"> <li>1. Magnification: 6:1 zoom, motorized</li> <li>2. Total Magnification               <ol style="list-style-type: none"> <li>a. around 4.0x to 24.5x with 10x eyepiece</li> <li>b. around 5.0x to 31.0x with 12.5x eyepiece</li> </ol> </li> <li>3. Focus Range: At least 75mm</li> <li>4. Objective/ Working Distance               <ol style="list-style-type: none"> <li>a. Working distance at least 200mm / focal length = 225 mm</li> </ol> </li> <li>5. Field of View: at least 51.4 – 8.6 mm Ø with 10x eyepiece</li> <li>6. Eyepieces               <ol style="list-style-type: none"> <li>a. Wide-field eyepieces for persons wearing glasses</li> <li>b. 10x dioptric adjustment</li> <li>c. ±5 diopter settings</li> <li>d. adjustable eyecup</li> </ol> </li> </ol>		

\_\_\_\_\_  
*Name and Signature of Authorized Representative*

## TECHNICAL SPECIFICATIONS

### SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<b><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</u></b> (continuation)	7. Direct Illumination with 2 LED Lamps <ul style="list-style-type: none"> <li>a. Main Light: Integrated LED illumination system for intensive uniform illumination of the field of view</li> <li>b. Continuously adjustable brightness with halogen-like color temperature</li> <li>c. 4 Coaxial Illumination: Illumination unit for generating a clear and stable Red Reflex, decreasing stray light through the sclera and increasing the image contrast</li> <li>d. Integrated keratoscope and slit illumination conversion filters allow surgeon to select preferred color temperature of main illumination</li> </ul> 8. Adjustable Coaxial illumination: diameter of coaxial illumination is adjustable between 4 mm and 23 mm via footswitch 9. Fine Focus: Available for assistant and integrated camera or external 1/3 camera with C-mount interface		

\_\_\_\_\_  
*Name and Signature of Authorized Representative*

## TECHNICAL SPECIFICATIONS

### **SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES**

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u><b>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</b></u> (continuation)	UPGRADEABILITY 1. Prepared for integration of video camera systems, digital recording and imaging systems and monitors 2. Connectors a. Numerous built-in connectors for video and control data transfer b. Internal power supply 12 to 24 VDC, and AC terminals 3. HD Video: Fully integrated 2D HD video recording  MANEUVERABILITY 1. Optics: a. 360° rotation b. 15° / + 105° motorized inclination tilt 2. XY Speed: Zoom linked XY speed 3. XY Range: 62 x 62mm 4. Balancing: Adjustable gas spring via balancing knob 5. Brakes: Floor stand with 4 electromagnetic brakes 6. Monitor Arm: 860 mm flexible arm with 4 axis for rotation and inclination, max. weight 15 kg and up to 32"		

*Name and Signature of Authorized Representative*

## TECHNICAL SPECIFICATIONS

### **SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES**

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u><b>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</b></u> (continuation)	<b>CONTROL</b> 1. Control unit <ul style="list-style-type: none"> <li>a. User-friendly, individually programmable touch-screen (up to 30 surgeons) for control of motor functions and light intensity</li> <li>b. Menu selection based on unique software for user-specific configuration</li> <li>c. Built-in electronic auto-diagnosis and user support</li> <li>d. Software independent hard keys and indicator for illumination</li> <li>e. Data shown by means of LCD</li> </ul> 2. Control Elements <ul style="list-style-type: none"> <li>a. Rotary handles</li> <li>b. 14 – Function wireless footswitch with optional back-up cable</li> </ul> 3. IR Sensor: Remote control of the HDR recorder 4. Indicators <ul style="list-style-type: none"> <li>a. LED for video record status</li> <li>b. Surgeon information panel for setting status</li> </ul> 5. Floor Stand: Four 360° rotating castors (Ø150 mm), parking brake 6. Materials <ul style="list-style-type: none"> <li>a. Coated with antimicrobial paint</li> <li>b. Conforming with RoHS1</li> </ul> 7. Load: Floor stand maximum 8.0 kg from microscope dove-tail ring interface 8. Weight: Floor stand approximate 350 kg without load		

*Name and Signature of Authorized Representative*

## TECHNICAL SPECIFICATIONS

### SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u><b>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</b></u> (continuation)	TECHNICAL DATA 1. Power Connection: 100 V to 240 V, 60 Hz 2. Protection Class: Class 1 3. Vitreo-Retinal Viewing System: motorized viewing system 4. INTRA- OPERATIVE OCT <ul style="list-style-type: none"> <li>a. B-Scan Length: at least 20 mm</li> <li>b. Scan Area: at least 314 mm<sup>2</sup></li> <li>c. Scan per seconds: 32 kHz</li> <li>d. Axial Resolution: 4 μm</li> <li>e. Imaging depth: 2.5 mm</li> <li>f. Wavelength: 860 mm</li> </ul> 5. Connected to a UPS device with at least 150% of equipment power capacity		

\_\_\_\_\_  
*Name and Signature of Authorized Representative*

## TECHNICAL SPECIFICATIONS

### SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<b><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</u></b> <i>(continuation)</i>	<ul style="list-style-type: none"> <li><b><u>ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</u></b></li> </ul> <p>Illumination</p> <ol style="list-style-type: none"> <li>Light source: LED via fiber optic</li> <li>Illumination Type: Stereo coaxial illumination with adjustable ratio between red reflex and field illumination</li> <li>Filters: Retina protection filter, Halogen Mode</li> </ol> <p>Optics</p> <ol style="list-style-type: none"> <li>Apochromatic</li> <li>Integrated assistant scope ( no external beam splitter)</li> <li>Zoom system (surgeon arm): Motorized continuous zoom</li> <li>Magnification range (surgeon arm): 3.5x - 21x (with f = 200 mm objective lens and 10x wide-field eyepieces)</li> <li>Focus system (surgeon arm): Motorized focus</li> <li>Focus range (surgeon arm): At least 70mm</li> <li>Magnification (assistant arm): Motorized continuous zoom</li> <li>Focus adjustment (assistant arm): Manual knob adjustment</li> </ol>		

*Name and Signature of Authorized Representative*

## TECHNICAL SPECIFICATIONS

### SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u><b>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</b></u> (continuation)	9. Eyepiece: 10X magnification 10. Tubes (surgeon arm): Tilttable tube with inverter function 11. Tubes (assist): Tilttable tube with inverter function 12. Objective lens: f=200mm  Footswitch 1. Connection: Capable of wired and wireless 2. Functions: 14 functions  Mechanical / Dimensions 1. XY coupling travel range: 60-65mm/60-65mm 2. Foot dimension: 800 - 810 x 800 – 810 mm 3. Stand height: not less than 1800mm 4. Microscope head tilt angle: +90° / - 20 ° 5. Suspension arm stroke: ±360mm 6. Microscope handles: with suspension arm lock release and programmable buttons		

*Name and Signature of Authorized Representative*



## TECHNICAL SPECIFICATIONS

### **SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES**

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<b><u>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</u></b> <i>(continuation)</i>	<p>Features / Accessories</p> <ol style="list-style-type: none"> <li>1. With Fundus viewing system option with automatic inverter and motorized function control</li> <li>2. Camera: 1-chip HD camera (does not require any beam splitter)</li> <li>3. Integrated overlay of surgical assistants for incision, LRI, rhexis, IOL alignment and microscope presets on surgeon oculars and video display</li> <li>4. Capability for data to be transferred from a compatible swept source biometry machine to the integrated overlay of surgical assistant to create overlays in the eyepiece for markerless IOL alignment</li> <li>5. Monitor and Video Recording: Integrated touchscreen monitor eye panel with built in PC with capability to capture high quality video and photos with the following specifications:               <ol style="list-style-type: none"> <li>5.1.1 Touchscreen: Projected Capacitive Touch (PCT) with anti-reflective coating and scratch proof</li> <li>5.1.2 Processor: Intel Core i5 6442EQ 1.9 Ghz or better</li> <li>5.1.3 Operating system: Windows 7</li> <li>5.1.4 Hard drive: SATA HDD 1 TB or higher for data</li> <li>5.1.5 Integrated color flat screen 24" or bigger</li> </ol> </li> </ol>		

*Name and Signature of Authorized Representative*

## TECHNICAL SPECIFICATIONS

### SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u><b>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</b></u> (continuation)	5.1.6 Video signals 5.1.6.1 PAL 576i50 5.1.6.2 NTSC 480i60 5.1.6.3 NTSC 1080i50 5.1.6.4 NTSC 1080i60 5.1.7 Ports 5.1.7.1 1 x CAN-Bus 5.1.7.2 2 x 1 Gigabit Ethernet 5.1.7.3 5 x USB 3.0 5.1.7.4 1 x potential equalization 5.1.8 Video input 5.1.8.1 1 x Y/C 5.1.8.2 1 x HD-SDI 5.1.9 Video output: 2 x HDMI 6. Connectivity: Capable of surgery data, import/export via USB (DICOM connectivity capable with a biometry machine) 7. With fast focus 8. With Programmable user presets		

*Name and Signature of Authorized Representative*

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

Item	Procuring Entity’s Specification	Bidder’s Specification as Technical Offer	Statement of Compliance
<p><u>SUPPLY,</u> <u>DELIVERY AND</u> <u>INSTALLATION OF</u> <u>ONE (1) UNIT</u> <u>SURGICAL</u> <u>MICROSCOPE</u> <u>WITH OCT</u> <u>CAPABILITY AND</u> <u>ONE (1) UNIT</u> <u>SURGICAL</u> <u>MICROSCOPE</u> <u>WITH SOFTWARE</u> <u>CONNECTIVITY</u> <u>TO OTHER</u> <u>OPHTHALMIC</u> <u>DIAGNOSTIC</u> <u>DEVICES</u> (continuation)</p>	<p>Fundus Viewing System</p> <ol style="list-style-type: none"><li>1. Inverter: Automatic motorized inverter</li><li>2. Focusing Range: At least 38mm, motorized</li><li>3. Lens Holder Rotation: 0° - 360° (in 30° increments)</li><li>4. Lenses: Aspheric, 60D &amp; 128D</li><li>5. Sterilization Tray: Metal tray</li><li>6. Focus: Motorized focus control</li></ol>		

Name and Signature of Authorized Representative

TECHNICAL SPECIFICATIONS

SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES

Name of Project

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<p><u>SUPPLY,</u> <u>DELIVERY AND</u> <u>INSTALLATION OF</u> <u>ONE (1) UNIT</u> <u>SURGICAL</u> <u>MICROSCOPE</u> <u>WITH OCT</u> <u>CAPABILITY AND</u> <u>ONE (1) UNIT</u> <u>SURGICAL</u> <u>MICROSCOPE</u> <u>WITH SOFTWARE</u> <u>CONNECTIVITY</u> <u>TO OTHER</u> <u>OPHTHALMIC</u> <u>DIAGNOSTIC</u> <u>DEVICES</u> (continuation)</p>	<p>Other terms &amp; conditions of acceptability:</p> <ol style="list-style-type: none"><li>1. The equipment must be BRAND-NEW and under warranty including spare parts for at least 2 years. Supplier must bind itself to conduct preventive maintenance on a quarterly basis at its own expense (including labor and cost of spare parts).</li><li>2. All machines must be manufactured by a known and reputable company compliant with ISO or its equivalent for equipment only.</li><li>3. The equipment / machine must be able to comply with its installation and operation qualification which would be conducted in the presence of SPMC Biomedical Technician, End-user and the Company Engineer.</li><li>4. In line with the ISO standard requirement, the supplier’s engineer or technical personnel must be capable of conducting annual calibration and the issuance of certificate for the said calibration based on the following data:<ol style="list-style-type: none"><li>a. Reference standards</li><li>b. National Institute of Standards and Technology traceability</li><li>c. Reference values</li><li>d. Validity period</li><li>e. Uncertainty of measurement</li></ol></li></ol>		
---	--	--	--

Name and Signature of Authorized Representative

## TECHNICAL SPECIFICATIONS

### **SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES**

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u><b>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</b></u> (continuation)	<p>5. Supplier must comply with the following conditions:</p> <ul style="list-style-type: none"> <li>a. Must be the exclusive or authorized distributor of the principal company of the equipment and the necessary consumables in the Philippines.</li> <li>b. Submit Certificate of Training of the Company Engineer / Technical Personnel and Product Specialist issued by the Principal or Manufacturer.</li> <li>c. Provide company response within 24 – 48 hours in case of technical problems or equipment breakdown.</li> <li>d. In case of machine downtime, the supplier is given three (3) days for remedial action (repair and or replacement of spare parts). On the 4th day, if the unit is not operational, a back-up machine shall be provided within the warranty period. . The supplier will be given fourteen (14) days to comply in providing back-up machine. A penalty will be imposed after the fourteenth day amounting to the average daily income of the machine. This is to cover for the income loss.</li> </ul> <p>6. The Supplier must submit the following:</p> <ul style="list-style-type: none"> <li>a. Three (3) copies of the Operating and Service Manuals in English during the delivery of the equipment.</li> <li>b. Notarized Certificate of availability of spare parts for the next ten (10) years.</li> </ul> <p>7. Cost of damages on the existing facility that will be incurred during delivery and installation of all equipment will be shouldered by the supplier.</p>		

*Name and Signature of Authorized Representative*

## TECHNICAL SPECIFICATIONS

### SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES

*Name of Project*

The bidder is required to complete its Statement of Compliance demonstrating how its items complied with the specification. Non-completion of Statement of Compliance for the item included in its bid shall be considered **non- responsive** for that item.

<i>Item</i>	<i>Procuring Entity's Specification</i>	<i>Bidder's Specification as Technical Offer</i>	<i>Statement of Compliance</i>
<u><b>SUPPLY, DELIVERY AND INSTALLATION OF ONE (1) UNIT SURGICAL MICROSCOPE WITH OCT CAPABILITY AND ONE (1) UNIT SURGICAL MICROSCOPE WITH SOFTWARE CONNECTIVITY TO OTHER OPHTHALMIC DIAGNOSTIC DEVICES</b></u> (continuation)	<p>8. The principal in coordination with the supplier must provide a comprehensive certification training program preferably onsite for the end-users and biomedical technician without additional cost to the procuring entity. Supplier must submit a <b>comprehensive training module</b> as part of the technical specification which covers product orientation, hands-on training and troubleshooting. The said training must be conducted by a certified product specialist. The module should indicate the following:</p> <ul style="list-style-type: none"> <li>- The objectives of the training</li> <li>- The methods or activities that shall be done during the training</li> <li>- The training outline that indicates the concepts and/or skills to be discussed/ taught.</li> <li>- The target participants of the training</li> <li>- The activity schedule indicating the no. of days or hours</li> </ul> <p>9. The Southern Philippines Medical Center has the right to terminate the contract for any violation in the terms and conditions stated in the technical specification and other reasons as stated in the General Conditions of the Contract (GCC).</p>		

The Bidder may submit offer which provides for superior specifications and/or better terms and conditions to the government at no extra cost. However, these shall not be given any bonus, credit or premium in the bid evaluation.

*Name and Signature of Authorized Representative*